



PC-1033

111th IAMM EQAS Microbiology: Bacteriology/ Serology
CMC MICRO EQAS
 Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
 Email: cgas@cmcvellore.ac.in, Twitter: @microcgas, Phone: +91-416-2282588



JULY 2022

111th EQAS EVALUATION REPORT

MEMBER ID:

M 1 2 1 2

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination

Peer group (n) = 229

Please refer the attached evaluation format for details on the criteria for evaluation

Marks Obtained: 6/10 (60%)

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score		
SE1	CRP	Negative	0.6	Negative	Not Applicable						2	2
SE2	CRP	Negative	0.6	Positive	31.0011	19.1529	6 to 192	1.6328	2	0		
SE3	CRP	Negative	0.6	Positive	20.3592	12.0561	6 to 192	1.0525	2	0		
								Not Applicable	2	2		
								Applicable	2	0		

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination

Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score		
SE1	RF	Negative	10	Negative	Not Applicable						2	2
SE2	RF	Negative	10	Negative	Not Applicable						2	2
SE3	RF	Negative	10	Negative	NOT EVALUATED*							

* As >30% participants scored <70% on this exercise, it has been excluded from the final evaluation.