



111th IAMM EQAS Microbiology: Bacteriology/ Serology
CMC MICRO EQAS
 Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
 Email: eqas@cmcvellore.ac.in, Twitter: @microeqas, Phone: +91-416-2282588

PC-1033

MEMBER ID:

| | | | | |
|---|---|---|---|---|
| M | 1 | 2 | 1 | 2 |
|---|---|---|---|---|

JULY 2022

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination

Peer group (n) = 229

Please refer the attached evaluation format for details on the criteria for evaluation

111th EQAS EVALUATION REPORT

Marks Obtained: 6/10 (60%)

| Parameter | Your Result | Your Value (mg/L) | Intended Result | Robust Mean | Robust SD | Range (mg/L) | Uncertainty of Assigned value | Z & Z' score | Max Marks | Your Score |
|-----------|-------------|-------------------|-----------------|-------------|-----------|--------------|-------------------------------|----------------|-----------|------------|
| SE1 | Negative | 0.6 | Negative | 31.0011 | 19.1529 | 6 to 192 | 1.6328 | Not Applicable | 2 | 2 |
| SE2 | Negative | 0.6 | Positive | 20.3592 | 12.0561 | 6 to 192 | 1.0525 | Not Applicable | 2 | 0 |
| SE3 | Negative | 0.6 | Positive | | | | | | 2 | 0 |

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination

Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

| Parameter | Your Result | Your Value (IU/mL) | Intended Result | Robust Mean | Robust SD | Range (IU/mL) | Uncertainty of Assigned value | Z & Z' score | Max Marks | Your Score |
|-----------|-------------|--------------------|-----------------|-------------|-----------|----------------|-------------------------------|--------------|-----------|------------|
| SE1 | Negative | 10 | Negative | | | Not Applicable | | | 2 | 2 |
| SE2 | Negative | 10 | Negative | | | | | | 2 | 2 |
| SE3 | Negative | 10 | Negative | | | | | | | |

NOT EVALUATED*

...70% on this exercise, it has been excluded from the final evaluation.

Disclaimer:


This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID: **M 1 2 1 2**


| SE1 | SE2 | SE3 | Marks obtained |
|-----|-----|-----|--------------------|
| 4 | 2 | 0 | 6 |
| 4 | 4 | 2 | Maximum marks = 10 |



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator



Dr. Rani Diana Sahni
Scientific Co-ordinator

Report Dispatch Date: 15.11.2022

***** End of Report *****