



Rapid Diagnostic Pvt. Ltd.



ISO 9001:2008  
CERTIFIED COMPANY

## Calibrator Calibration

<b>Instrument:</b> 5 Part Haematology Analyzer	<b>Model:</b> BC -6200	<b>Serial No.:</b> TW-02000911	<b>Date of Cal.</b> 14.07.2022			
<b>Mode: OV-WB</b>	<b>Calibrator Lot No. PLUS0722</b>			<b>Exp.Date:05.08.2022</b>		
<b>Parameter</b>	<b>WBC</b>	<b>RBC</b>	<b>HGB</b>	<b>MCV</b>	<b>PLT</b>	<b>WBC-D</b>
<b>TARGET</b>	9.05	4.59	13.6	92.5	255	9.05
1.	8.82	4.78	14.2	92.9	236	7.95
2.	8.69	4.66	14.2	92.5	237	8.35
3.	8.88	4.69	13.9	92.3	242	8.32
4.	8.92	4.69	14.2	92.6	239	8.44
5.	8.95	4.74	13.9	92.1	245	8.35
6.	8.84	4.65	14.3	92.1	229	8.37
<b>Mean</b>	8.87	4.63	14.0	92.3	234	8.35
<b>CV( %)</b>	1.485	1.089	0.58	0.33	2.8	2.465
<b>New Cal. Factor (%)</b>	103.25	99.45	97.98	101.28	102.42	
<b>Old Cal. Factors (%)</b>	103.20	93.96	96.86	100.49	96.75	
<b>Old TC1 (%)</b>						101.15
<b>New TC1 (%)</b>						106.41

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju  
Service Engineer

N. Mahender Reddy  
Zonal Service Manager



Rapid Diagnostic Pvt. Ltd.



## Calibration Certificate

Customer Name	Telangana Diagnostics VIKARABAD
Address	Area Hospital VIKARABAD, VIKARABAD-508213
Contact Person	Mr. SRINU
Contact No.	9110375214
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-02000911
Instrument Status	Under Warranty
Calibration Done On	14.07.2022
Next Calibration Due On	13.07.2023

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju  
Service Engineer

N. Mahender Reddy  
Zonal Service Manager