

CALIBRATION REPORT
Centrifuges, Tabletop (Centrifuge)



Report No:	CAL-MZ-2023-02-0668	Date: 31-01-2023	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	Thingsuthliah CHC		
State	Mizoram	District	Aizawl East

Medical Equipment Details			
ME Number	MZZ1AWLPH0600048	Manufacturer	Remi
ME Name	Centrifuges, Tabletop (Centrifuge)	Model	NA
Department	Laboratory	Serial No	NA-MZZ1AWLPH0600048
Calibrated at	On Site	Condition of ME	Visually Good
Date of Calibration	30-01-2023	Due for Calibration	29-01-2024

Environmental Conditions			
Temperature in °C	22	Humidity in % Rh	56

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	TRI-TE-0014	Digital Tachometer	Amprobe	Tach-10	16060040	06-09-2023

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	rpm	1500	1522	1425 to 1575	Pass
2	Speed	rpm	2500	2577	2375 to 2625	Pass
3	Speed	rpm	3500	3547	3325 to 3675	Pass

CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval.

Calibration Status

Passed

Failed

Shoib
Calibrated By
Shoib Malik

S. Senthil
Authorised Signatory
Senthil S