

**CALIBRATION REPORT**

Centrifuge Table top



Report No:	CAL-CG-2022-11-0507	Date: 02-12-22	Page 01 of 02
------------	---------------------	----------------	---------------

Customer Details			
Health Facility	Korba DH		
State	Chhattisgarh	District	Korba

Medical Equipment Details			
ME Number	CGZ3KOBBDH0170009	Manufacturer	NA
ME Name	Centrifuges, Tabletop(Centrifuge)	Model	Na
Department	Lab	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	12-11-22	Due for Calibration	11-11-23


Environmental Conditions			
Temperature in °C	32°C	Humidity in % Rh	39%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0104	Digital Tachometer	Amprobe	TACH-10	18080001	22-7-2023

Performance Results						
No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	Speed	rpm	1500	1500	1425 to 1575	PASS
2	Speed	rpm	2500	2605	2375 to 2625	PASS
3	Speed	rpm	3500	3645	3325 to 3675	PASS

**CALIBRATION REPORT**  
Centrifuge Table top



<b>Conditions of Issue of report</b>	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval.	
<b>Calibration Status</b>	Passed <input checked="" type="checkbox"/> Failed
<b>Calibrated By</b>	<b>Authorised Signatory</b>
Suman Mahant	 A.Raja