

# CALIBRATION REPORT

Rotators(Lab Rotator)



Report No:	CAL-CG-2022-11-0507	Date: 28-1-23	Page 01 of 01
------------	---------------------	---------------	---------------

Customer Details			
Health Facility	Korba		
State	Chhattisgarh	District	Korba DH

Medical Equipment Details			
ME Number		Manufacturer	Remi
ME Name	Rotators(Lab Rotator)	MOdel	Rs-12R
Department	L <sub>3</sub> b	Serial No	ZEKS-38616
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	28-1-23	Due for Calibration	27-1-24


Environmental Conditions			
Temperature in °C	32°C	Humidity in % Rh	39%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
	CHC/TE/0104	Digital Tachometer	Amprobe	TACH-10	18080001	22-7-2023

Performance Test						
No	Parameters	Unit	Set	Measured	Acceptable Range	Remarks
1	Speed	rpm	80	81	78 to 82	Pass
2	Speed	rpm	100	102	98 to 102	Pass

Conditions of Issue of report	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval.	

Calibration Status	Passed	Failed O
Calibrated By	Authorised	
A.Raja	A.Raja	



Freez-2

# CALIBRATION REPORT

Refrigerators, Laboratory (Lab refrigerator)



**TRIMED**  
SOLUTIONS (I) PVT. LTD

Report No: | **CAL-CG-2022-11-0542** | Date: 28-1-23 | Page 01 of 02

Customer Details			
Health Facility	Korba DH		
State	Chhattisgarh	District	Korba

Medical Equipment Details			
ME Number	-	Manufacturer	Samsung
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	RT34M55S8/H1/2019
Department	Lab	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	28-1-23	Due for Calibration	27-1-24

Environmental Conditions	
Temperature in °C	32°C   Humidity in % Rh 139%



Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0089	THERMOMETER	METRAVI	DMT-900	171112536	08-5-2023

Performance Results				
No	Parameter	Measured Value	Acceptable Range	Remarks
1	Temperature	4.5	3.5 to 4.5°C	Pass

# CALIBRATION REPORT



Refrigerators, Laboratory (Lab refrigerator)

<b>Conditions of Issue of report</b>	
1. The Standards used are traceable to National standards.	
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.	
<hr/>	
<b>Calibration Status</b>	Passed <input checked="" type="checkbox"/> Failed <input type="checkbox"/>
<b>Calibrated By</b>	<b>Authorised Signatory</b>
A.Raja	  A.Raja