

Freeze - 2

CALIBRATION REPORT



Refrigerators, Laboratory (Lab refrigerator)

Report No:	CAL-CG-2022-11-0535	Date: 02-12-22	Page 01 of 02
------------	---------------------	----------------	---------------

Customer Details			
Health Facility	Korba DH		
State	Chhattisgarh	District	Korba

Medical Equipment Details			
ME Number	CGZ3KOBHDH0170377	Manufacturer	Godrej
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	NA
Department	Lab	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	12-11-22	Due for Calibration	11-5-23

Environmental Conditions			
Temperature in °C	32°C	Humidity in % Rh	39%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0082	THERMOMETER	METRAVI	DMT-900	171112532	02-12-2022

Performance Results				
No	Parameter	Measured Value	Acceptable Range	Remarks
1	Temperature	3.7	3.5 to 4.5°C	Pass

CALIBRATION REPORT



Refrigerators, Laboratory (Lab refrigerator)

Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status

Passed

Failed

Calibrated By

Priya Rajwade

Authorised Signatory



A.Raja