

Freez-2

CALIBRATION REPORT

Refrigerators, Laboratory (Lab refrigerator)



TRIMED
SOLUTIONS (I) PVT. LTD

Report No: | **CAL-CG-2022-11-0542** | Date: 28-1-23 | Page 01 of 02

Customer Details			
Health Facility	Korba DH		
State	Chhattisgarh	District	Korba

Medical Equipment Details			
ME Number	-	Manufacturer	Samsung
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	RT34M55S8/H1/2019
Department	Lab	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	28-1-23	Due for Calibration	27-1-24

Environmental Conditions	
Temperature in °C	32°C
Humidity in % Rh	139%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0089	THERMOMETER	METRAVI	DMT-900	171112536	08-5-2023

Performance Results				
No	Parameter	Measured Value	Acceptable Range	Remarks
1	Temperature	4.5	3.5 to 4.5°C	Pass

CALIBRATION REPORT



TRIMED

soumoNs (I) ***** LTD

Refrigerators, Laboratory (Lab refrigerator)

Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status


Passed

Failed

Calibrated By

A.Raja

Authorised Signatory



A.Raja