



Calibrator Calibration

| | | | | | | |
|---|------------|------------------------------------|------------|-----------------------------------|----------------------------|------------------------------------|
| Instrument: 5 Part Haematology Analyzer | | Model: BC -6200 | | Serial No.: TW-9C000818 | | Date of Cal.: 02.12.2022 |
| Mode: OV-WB | | Calibrator Lot No. PLUS1122 | | | Exp.Date:05.12.2022 | |
| Parameter | WBC | RBC | HGB | MCV | PLT | WBC-D |
| TARGET | 9.16 | 4.56 | 13.4 | 91.8 | 250 | 9.16 |
| 1. | 8.99 | 4.65 | 14.2 | 88.7 | 234 | 8.72 |
| 2. | 8.86 | 4.62 | 14.2 | 88.9 | 242 | 8.87 |
| 3. | 8.96 | 4.58 | 14.1 | 88.7 | 231 | 8.37 |
| 4. | 9.14 | 4.61 | 14.1 | 88.4 | 236 | 8.90 |
| 5. | 8.98 | 4.59 | 14.1 | 88.5 | 242 | 8.62 |
| 6. | 9.21 | 4.62 | 14.2 | 87.9 | 241 | 8.79 |
| Mean | 9.02 | 4.61 | 14.1 | 88.5 | 238 | 8.71 |
| CV(%) | 1.420 | 0.546 | 0.40 | 0.36 | 2.1 | 2.251 |
| New Cal. Factor (%) | 101.51 | 98.87 | 94.86 | 103.70 | 105.20 | |
| Old Cal. Factors (%) | 103.1 | 100.54 | 94.25 | 105.0 | 103.97 | |
| Old TC1 (%) | | | | | | 101.48 |
| New TC1 (%) | | | | | | 103.59 |

For Rapid Diagnostic Pvt. Ltd.


N. Ramakrishna Reddy
 Service Engineer


N. Mahender Reddy
 Zonal Service Manager



Calibration Certificate

| | |
|-------------------------|--|
| Customer Name | Telangana Diagnostic Hub Nizambad |
| Address | Govt hospital, Nizambad-505301 |
| Contact Person | Dr. Divya |
| Contact No. | 8008244744 |
| Instrument | 5 Part Haematology Analyzer |
| Model | BC 6200 |
| Serial No. | TW-9C000818 |
| Instrument Status | Under Warranty |
| Calibration Done On | 02.12.2022 |
| Next Calibration Due On | 01.12.2023 |

For Rapid Diagnostic Pvt. Ltd.



N. Ramakrishna Reddy
Service Engineer



N. Mahender Reddy
Zonal Service Manager