

Customer Name	THK KATTAPPANA	Certificate No.	THK018
Address	Kattappana, Kerala 685515	Date of Calibration	18/10/2022
		Valid Upto	17/10/2023
		Date of Issue	19/10/2022

DUC* Details					
DUC* Name	Make/Model No	S No.	ASSET ID	LOCATION	Visual Inspection
VDRL Rotator	ROTEK/PVD R01D	204	620555	LABORATORY	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	ELECTRICAL SAFETY ANALYZER	THERMO HYGROMETER
Make	FLUKE	HTC-2
Model/SR No.	ESA 612/2959012	...
Equipment Id	CHPL/MT-KL/ESA/002	
Calibration certificate No	ET/220124A/01	UCSL/07-22/184-01
Calibration Validity	25/01/2023	20/07/2023
Calibration By	STIC	UNIQUE CALIBRATION SOLUTIONS

**STANDARD USED IS IEC 60601-1

ELECTRICAL SAFETY TEST VALUES

TEST NO.	NAME OF THE TEST	READING	HIGH LIMIT	LOW LIMIT	UNIT	STATUS
1	MAIN VOLTAGE -LIVE TO NEUTRAL	228.6	250	225	V	PASS
2	MAIN VOLTAGE -LIVE TO EARTH	228.5	240	220	V	PASS
3	MAIN VOLTAGE-NEUTRAL TO EARTH	0.3	3	0	V	PASS
4	EQUIPMENT CURRENT	0.2	16	0.1	mA	PASS
5	PROTECTIVE EARTH RESISTANCE	0.3	2	0	Ω	PASS
6	EARTH LEAKAGE CURRENT - NC	0.6	1	0.5	mA	PASS
7	EARTH LEAKAGE CURRENT - OC	0.5	1	0.5	mA	PASS
8	ENCLOSURE LEAKAGE CURRENT -NC	0.4	0.5	0.1	mA	PASS
9	ENCLOSURE LEAKAGE CURRENT -OC	0.5	0.5	0.1	mA	PASS
10	INSULATION RESITANCE					
	1.MAINS-PROTECTIVE EARTH	OL	100	0	M Ω	PASS
	2.APPLIED PART TO PROTECTIVE EARTH	NA	100	0	M Ω	PASS
	3.MAINS TO APPLIED PART	NA	100	0	M Ω	PASS
	4.MAINS TO NEUTRAL	OL	100	0	M Ω	PASS
	5.APPLIED PART TO NEUTRAL	NA	100	0	M Ω	PASS

*NOTE (NC-NORMAL CONDITION, OC-OPEN CONDTION). REMARKS- *DUC- DEVICE UNDER CALIBRATION

CALIBRATED BY

ACHUTH R K



ISSUED AND APPROVED BY

PREJITH T B
TECHNICAL MANAGER

