



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



CALIBRATION CERTIFICATE

Customer name And Address	M/S: SRV Path Labs Pvt, Ltd. B-2/ 28A, Yamuna Bihar, Delhi 110053 (Near Bhajan Pura Petrol Pump)	ULR No.	CC223922000001703 F
		Service request No & date	P/02/27/04/2022
		Date of calibration	27/04/2022
		Valid up to	26/04/2023
		Date of issue	28/04/2022

Instrument Details

Instrument name	Digital Thermometer	I.D. No.	DTM-03
Make / Model No.	Haier	Location	Refrigerator
Range / Size	-20 to 100 °C	Accuracy	-----
Least Count	0.1 °C	Visual Inspection	OK

Detail of reference standards & Major equipments used

Equipment Name	RTD With Digital Indicator		
Make	Tempens		
Model / SR No.	11080368		
Certificate No.	CC325722000000135F		
Calibration Validity	10/02/2023		
Calibration by	KL Technology		

Environmental Condition	Temperature	25±4 °C	Calibration Reference	IS:6274
	Relative Humidity	50±20 %	Calibration Procedure	CP-02T

Calibration Results


Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	-20.17	-20.3	±0.6 °C
02.	-15.13	-14.5	
03.	-10.11	-9.7	
04.	0.3	0.5	

Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%.
- ❖ (3) The above results are valid at the time of and under the stated conditions measurement.
- (4) This certificate is refers only to the particular item submitted for calibration.
- (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration Engg./TM)
(D.B SAXSENA)
Form No.-QF-47

Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1



End of Report---

