



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

ALL ANALYSER SUMMARY REPORT - SEPTEMBER 2022



PC-1024

Lab Name	REDCLIFFE LIFETECH PVT LTD	Lab No	15553
Constituent Group	HbA1c	Date of Result Entered	20/09/2022
PT item	Lyophilized Whole blood	Date of Report Published	05/10/2022

Sl.No	Constituent Name	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Biorad (D10 / D100 / Variant Turbo)	644	5.76	4.13	0.24			0.02
			Tosoh G8 - HPLC / HPLC-723G8)	143	5.53	5.16	0.28			0.05
			Any Analyser	111	5.35	10.08	0.54	5.3 %	-0.09	0.10

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS:

REDCLIFFE LIFETECH PVT LTD
 6 A/F-02, VRINDAVAN YOJNA, RB ROAD, OPP AKASH ENCLAVE BABHNAN BASTI
 LUCKNOW
 UTTAR PRADESH, 226021

Coordinator Contact details:
 Email: clinq@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****