


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 221016A/N-102	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 16 / 10 / 2022	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 15 / 04 / 2023	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No. &amp; DATE</b>	: 221016A/N Dt. 16.10.2022	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC342722000000661F Dt. 18.10.2022	<b>Cal. At</b>	: Site / Lab

**CALIBRATED FOR** : M/s Redcliffe Lifetech Pvt. Ltd.  
 Umeed Rehabilitation Centre 2nd Floor F-01,  
 Sector-6A, Vrindavan Yojna, R.B Road, Lucknow.

<b>ITEM CALIBRATED</b>	: Micro Pipette		
<b>Make</b>	: Fine Care Corporation	<b>Range</b>	: (10 to 100) µl
<b>L. Count</b>	:	<b>ID No.</b>	: RCL/BIO/PIP-02
<b>Location</b>	: Clinical Biochemistry	<b>S. No.</b>	: QB580166

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 3°C } R.H. { 40% to 60% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		10	10.09	(-) 0.09
100		50	50.22	(-) 0.22
		100	100.36	(-) 0.36

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.6 µl at coverage factor k=2.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

