SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27, Plot No.151, Street No.3, Teacher's Colony, East Marredpally, Secunderabad - 26, Telangana

Tel: 040 2773 2341/342, 2773 1510, Mob: 77299 91231/33, 98480 46524, E-mail_simco hyd@gmail.com, Info@simco-india.com_www.simcocal.brationlaboratory.com





CC-2806

CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

AN ISO 9001 : 2015

CERTIFIED COMPANY

Page: 1 of 1

Certificate No.: SL2212PS0456-002	Issue Date : 31-12-2022	
1. Customer Name & Address:		
M/S. Telangana Diagnostics,	ULR - CC 2 8 0 6 2 2 1 0 0 0 0 4 5 6 3 F	
Telangana Diagnostics Hub ,	Reference Date : 29-12-2022	
Rajanna Sircilla, Opp MRO Office Vidya Nagar - 502 103.	Calibration Date : 30-12-2022	
	Calibration Due Date : 29-12-2023	

2. Details of Unit Under Calibration:

Description	:	Centrifuge			
Make	:	REMI	SI No	: ZGKN-26043	
Range		Upto to 6000 RPM			
Resolution	- 10	10 RPM			
Model No	:	R-8C plus			

3. Detail of Standard Instruments Used :

Instrument Used	SI / Id No	Valid up to	Certificate No.
Digital Tachometer	175-0010V	30-03-2023	C-220326-7-1

4. Environmental Conditions: Standard Temperature: (24±4)°C

5. Calibration Procedure: SOP-PE-08

6. Mechanical Calibration: Speed

7. Calibration Results:

S. No.	Standard Reading (RPM)	UUC Reading (RPM)	Error (RPM)	Expanded Uncertainty in (± RPM)	
1	499.7	500	0.3	6	
2	999.1	1000	0.9	6	
3	1998.5	2000	1.5	6	
4	2997.8	3000	2.2	6	
5	4995.8	5000	4.2	6	

8. Remarks:

- The instrument/equipment is in good condition and was calibrated at Site
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions. C
- The calibration interval is determined based on customer's requirements.
- e The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- h The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Expanded Uncertainty is reported approximately at 95.45% confidence level with coverage factor k = 2.

K. B Bross Calibrated by

N.V. Kameswara Rao Technical Manager Authorised Signatory

Relative Humidity: (50±20) % RH