

Telangana Diagnostics

Department of Health & Family Welfare
Government of Telangana



RAJANNA SIRCILLA DISTRICT

Department of Microbiology

The PT report which we have shared in the application that report is dispatched on 15-11-2022 and that is the latest report we have in the last 6 months. We have sent December results to CMC Micro EQAS and we are waiting for their report. We will share that report to you as soon as we receive it. Meanwhile, please consider the November 2022 report as the latest.

Thanks and Regards
TD-Hub-Rajanna Sircilla



PC-1033

111th IANM EQAS Microbiology: Bacteriology/ Serology
CMC MICRO EQAS

Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
 Email: cgas@cmcvellore.ac.in, Twitter: @microeqas, Phone: +91-416-2282588



JULY 2022

111th EQAS EVALUATION REPORT

MEMBER ID:

M 1 2 1 1

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination

Peer group (n) = 229

Please refer the attached evaluation format for details on the criteria for evaluation

Marks Obtained: 10/10 (100%)

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score		
SE1	CRP	Negative	Negative	Not Applicable							2	2
SE2	CRP	Positive	Positive	31.0011	19.1529	6 to 192	1.6328	0.9	2	2		
SE3	CRP	Positive	Positive	20.3592	12.0561	6 to 192	1.0525	0.3	2	2		

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination

Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score		
SE1	RF	Negative	Negative	Not Applicable							2	2
SE2	RF	Negative	Negative	Not Applicable							2	2
SE3	RF	Positive	Negative	NOT EVALUATED*								

*As >30% participants scored <70% on this exercise, it has been excluded from the final evaluation.

Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 1 2 1 1


SE1	SE2	SE3	Marks obtained	
4	4	2	10	100%
4	4	2	Maximum marks = 10	



Dr. Rani Diana Sahni
Scientific Co-ordinator



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator

Report Dispatch Date: 15.11.2022

***** End of Report *****