

VIEW LAB MONTHLY SUMMARY

Lab Name	REGIONAL RESEARCH INSTITUTE OF UNANI MEDICINE	Details About Robust Analysis
Lab No	0151	Detail About Monthly Summary
Month	January	Detail about SDI
Year	2023	
Constituent Group	HbA1c	

[Click on the analyte to view Graphical Data](#)

All Analyser Result
Print

Date of Result Entered : **11/01/2023**

Date of Report Published : **09/02/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Any Analyser	311	6.09	11.39	0.69	6.7 %	0.88	0.08

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out