

VIBRANT Dx,

3-A, KUMARAN STREET, BALAMBIGAI FLATS, KEELKATTALAI, CHENNAI-600 117

9360157929

REF.# VDX/RRIU/SAA-CALIB/001

DATE : 12-Oct-2021

**Regional Research Institute of Unani Medicine,
West Madha Church Road,
Royapuram,
Chennai-600 013**

Dear Sir,

As desired by you please find the Calibration report of ERBA CHEM-7

Serial No.090650

1. Year of manufacturing : 2009.
2. Range of the instrument : Optical density Linear form 0-2.5
3. Date of Calibration Check: ____12-Oct-2021____
4. Date of Next Calibration Check Due: ____11-Oct-2022____

VIBRANT Dx,

3-A, KUMARAN STREET, BALAMBIGAI FLATS, KEELKATTALAI, CHENNAI-600 117

9360157929

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	OK / Not OK
Logic 5 V	5.00V	OK
Line voltage	221V	OK
Lamp Voltage 12v	11.97V	OK
Temperature 25°C	25°C	OK
Temperature 30°C	30°C	OK
Temperature 37°C	37°C	OK

VIBRANT Dx,

3-A, KUMARAN STREET, BALAMBIGAI FLATS,KEELKATTALAI,CHENNAI-600 117


9360157929

Sr.No.	Filters	Pre Channel	Status
1	340	61 K	OK
2	405	43 K	OK
3	450	64 K	OK
4	505	53 K	OK
5	546	35 K	OK
6	578	107 K	OK
7	600	90 K	OK
8	670	101 K	OK

Thanking you and assuring of our prompt services at all times.

Yours Sincerely,

For Vibrant Dx.,



S.MURLIDHARAN
Technical Service Dept.,

PLACE: CHENNAI-117

DATE: 12-Oct-2021