

VIBRANT Dx,

3-A, KUMARAN STREET, BALAMBIGAI FLATS, KEELKATTALAI, CHENNAI-600 117

9360157929

REF.# VDX/RRIU/SAA-CALIB/002

DATE : 26-Oct-2021

**Regional Research Institute of Unani Medicine,
West Madha Church Road,
Royapuram,
Chennai-600 013**

Dear Sir,

As desired by you please find the Calibration report of ROBONIK PRIETEST EASYLAB

Serial No.PEL0371113RBK

1. Year of manufacturing : 2013.
2. Range of the instrument : Optical density Linear form 0-2.5
3. Date of Calibration Check: 29-Oct-2021
4. Date of Next Calibration Check Due: 28-Oct-2022

VIBRANT Dx,

3-A, KUMARAN STREET, BALAMBIGAI FLATS, KEELKATTALAI, CHENNAI-600 117

9360157929

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	OK / Not OK
Logic 5 V	5.00V	OK
Line voltage	221V	OK
Lamp Voltage 12v	5.99 V	OK
Temperature 25°C	25°C	OK
Temperature 30°C	30°C	OK
Temperature 37°C	37°C	OK

VIBRANT Dx,

3-A, KUMARAN STREET, BALAMBIGAI FLATS, KEELKATTALAI, CHENNAI-600 117


9360157929

Sr.No.	Filters	Pre Channel	Status
1	340	3.3	OK
2	405	3.8	OK
3	510	3.8	OK
4	545	3.9	OK
5	578	4.6	OK
6	630	4.8	OK

Thanking you and assuring of our prompt services at all times.

Yours Sincerely,

For Vibrant Dx.,



S.MURLIDHARAN
Technical Service Dept.,

PLACE: CHENNAI-117

DATE: 26-Oct-2021