



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 1979

Distribution No.: 157-E

Month/Year: August/2022

Instrument ID: BC-6200

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 22-10-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	7.79	7.68	15.47	17.38	0.1220	-0.61	0.11	0.17	0.0110	-0.40
RBC x10 ⁶ /µl	1	5.36	5.35	10.71	10.98	0.0110	-0.88	0.01	0.04	0.0030	-0.67
Hb g/dl	1	10.3	10.2	20.5	21.1	0.0220	-1.01	0.1	0.1	0.0070	0.00
HCT%	1	44.4	44.2	88.6	71.3	0.1690	3.78	0.2	0.3	0.0230	-0.27
MCV-fl	1	82.8	82.5	165.3	129.15	0.2640	4.85	0.3	0.2	0.0120	0.45
MCH-Pg	1	19.1	19.1	38.2	38.5	0.0510	-0.25	0	0.1	0.0090	-0.79
MCHC-g/dl	1	231	230	461	59.2	0.1610	90.34	1	0.2	0.0130	3.60
Plt. x10 ³ /µl	1	234	231	465	465	2.38	0.00	3	7	0.49	-0.49
Retic %	2	1.8		1.8							

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs= , Poly=12 L=11, E=1, Mono/Promono=0 , B1=07 P.M.=06, Mye=16, Meta=24, Other=BAND FORMS -18	Poly: 19 - 36, Myelo: 20 - 40, Meta: 9 - 20, Lympho: 2 - 6, Promyelo: 1 - 6, nRBC/ Baso/ Eos/ Mono /Blast: 0 - 5		
RBC Morphology	3	Normocytic hypochromic RBCs admixed with microcytic hypochromic,nRBCs seen (3/100) Anisocytosis seen.	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis		
Diagnosis	3	Chronic myeloid Leukemia (Chronic phase)	Chronic Myeloid Leukemia (Chronic Phase)		

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 157--E	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	313	311	85.85	89.07	10.29	4.18	3.86	6.75
RBC x10⁶/µl	1	313	313	87.86	90.42	5.43	3.83	6.71	5.75
Hb g/dl	1	313	313	85.94	90.73	5.11	4.47	8.95	4.8
HCT%	1	313	310	92.26	89.03	4.52	4.52	3.22	6.45
MCV-fl	1	313	310	91.94	88.39	3.87	9.03	4.19	2.58
MCH-Pg	1	313	309	85.76	91.91	5.5	4.21	8.74	3.88
MCHC-g/dl	1	313	309	91.59	91.91	5.18	4.53	3.23	3.56
Plt. x10³/µl	1	313	311	94.53	87.14	4.5	6.43	0.97	6.43
ReticCount%	2	313	288	97.92	93.4	1.39	1.04	0.69	5.56
PS Assessment	3	313	289	Satisfactory :74.77%, Borderline Sat. :9.58%, Unsatisfactory :15.65%					

***Comments:**

1). Among Lab (EQA) : CBC result for HCT, MCV & MCHC unacceptable, please check calibration/human error.Remaining results acceptable.

2). Within Lab (IQA) : MCHC & RETIC result is unacceptable, please check precision/human error.Remaining precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme’s website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



TITLE

EQAS CORRECTIVE ACTION FORM

Issue No. 01

Page 1 of 1

EQAS Details	AIMS PATHOLOGY
Analyte:	HCT
Month:	AUGUST
Date Sample Tested:	08-09-2022

SPECIMEN HANDLING		
Were specimens received in an acceptable condition?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were specimens stored according to the instructions on the result forms?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the samples hemolyzed?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Were samples tested within the time allowed for sample stability?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
If applicable, were the samples reconstituted correctly?	Yes <input type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
CLERICAL ERRORS		
Were the results transcribed onto the result forms correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results transcribed from the result forms to the website correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded on the correct result form?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Was the correct instrument/reagent/kit selected?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded in the correct units?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results on your evaluation the same as the results you reported?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
QUALITY CONTROL		
Were quality control materials within the acceptable range on the date of PT testing? (Verify the quality control acceptable range in use.)	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Is there any indication of trending or shifting of the control results?	Yes <input type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
CALIBRATION		
Were there any problems with the most recent calibration?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
When was the last calibration performed?		
How often is a calibration performed?		
When was the last calibration verification performed?		
Notes: _____		

INSTRUMENT		
Were instrument problems noted the day the samples were tested?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Has there been any recent maintenance on the analyzer?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>

PREPARED & REVIEWED BY : CONSULTANT PATHOLOGIST: DR. B. JYOTHI	APPROVED & ISSUED BY: LAB HEAD: DR. B. JYOTHI
<i>B. Jyothi</i>	<i>B. Jyothi</i>



TELANGANA DIAGNOSTICS

Form: TD/QSP/08-EQCAR

TITLE

EQAS CORRECTIVE ACTION FORM

Issue No. 01

Page 1 of 1

Have you contacted your analyzer manufacturer for assistance?

Yes No

Notes:

REAGENTS

Were the reagents stored properly?

Yes No

Were the reagents expired or was the open vial stability exceeded?

Yes No

Have there been any changes in reagent manufacturer or formulation?

Yes No

Notes:

TESTING PERSONNEL

Date of last competency assessment for testing personnel

Yes No

Review assay procedure and proficiency test sample preparation instructions with testing personnel to ensure that instructions were followed

Yes No

Review with testing personnel how samples were loaded to rule out misidentification or transposition of samples.

Yes No

Notes:

Corrective Action:

It is a random error

Person Performing Investigation: G. Sateesh Kumar

Date: 16/10/2022

Lab Director: Dr. B. Jyothi

Date: 16/10/2022

PREPARED & REVIEWED BY :
CONSULTANT PATHLOGIST: DR.B.JYOTHI

B. Jyothi

APPROVED & ISSUED BY:
LAB HEAD: DR.B.JYOTHI

B. Jyothi

CONTROLLED COPY



TITLE

EQAS CORRECTIVE ACTION FORM

Issue No. 01

Page 1 of 1

INVESTIGATION SUMMARY: ROOT CAUSE

Pre-analytic Phase of Testing	Analytic Phase of Testing	Post-Analytic Phase of Testing
<input type="checkbox"/> PROBLEM WITH PT SAMPLE <input type="checkbox"/> SAMPLE PROCESSING <input type="checkbox"/> DATA ENTRY <input type="checkbox"/> OTHER (SPECIFY): _____	<input type="checkbox"/> METHODOLOGICAL PROBLEM <input type="checkbox"/> TECHNICAL PROBLEM <input type="checkbox"/> REAGENT PROBLEM <input type="checkbox"/> CALIBRATOR PROBLEM <input type="checkbox"/> OTHER (SPECIFY): _____	<input type="checkbox"/> CLERICAL ERROR <input type="checkbox"/> REPORTING PROBLEM <input type="checkbox"/> NO EXPLANATION AFTER INVESTIGATION <input type="checkbox"/> OTHER (SPECIFY): _____

PREVENTION

Preventive action proposed

We will monitor performance of HCT Parameter closely

Preventive action Plan

we will monitor performance of HCT Parameter in next cycle of EQAS

Responsibility

Empty box for responsibility

Date	16/10/2022	Testing Personnel	Gr. Sateesh Kumar
Date	16/10/2022	Department Technical In charge	Dr. B. Jyothi

PREPARED & REVIEWED BY: CONSULTANT PATHOLOGIST: DR.B.JYOTHI	APPROVED & ISSUED BY: LAB HEAD: DR.B.JYOTHI



TITLE

EQAS CORRECTIVE ACTION FORM

EQAS Details	AIIMS PATHOLOGY
Analyte:	MCHC
Month:	AUGUST
Date Sample Tested:	08-09-2022

SPECIMEN HANDLING		
Were specimens received in an acceptable condition?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were specimens stored according to the instructions on the result forms?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the samples hemolyzed?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Were samples tested within the time allowed for sample stability?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
If applicable, were the samples reconstituted correctly?	Yes <input type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
CLERICAL ERRORS		
Were the results transcribed onto the result forms correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results transcribed from the result forms to the website correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded on the correct result form?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Was the correct instrument/reagent/kit selected?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded in the correct units?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results on your evaluation the same as the results you reported?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
QUALITY CONTROL		
Were quality control materials within the acceptable range on the date of PT testing? (Verify the quality control acceptable range in use.)	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Is there any indication of trending or shifting of the control results?	Yes <input type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
CALIBRATION		
Were there any problems with the most recent calibration?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
When was the last calibration performed?		
How often is a calibration performed?		
When was the last calibration verification performed?		
Notes: _____		

INSTRUMENT		
Were instrument problems noted the day the samples were tested?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Has there been any recent maintenance on the analyzer?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>

PREPARED & REVIEWED BY : CONSULTANT PATHOLOGIST: DR.B.JYOTHI	APPROVED & ISSUED BY: LAB HEAD: DR.B.JYOTHI
<i>B. Jyothi</i>	<i>B. Jyothi</i>



TELANGANA DIAGNOSTICS

Form: TD/QSP/08-EQC

TITLE

EQAS CORRECTIVE ACTION FORM

Issue No

Page 1 of 1

Have you contacted your analyzer manufacturer for assistance?

Yes No

Notes: _____

REAGENTS

Were the reagents stored properly?

Yes No

Were the reagents expired or was the open vial stability exceeded?

Yes No

Have there been any changes in reagent manufacturer or formulation?

Yes No

Notes: _____

TESTING PERSONNEL

Date of last competency assessment for testing personnel

Yes No

Review assay procedure and proficiency test sample preparation instructions with testing personnel to ensure that instructions were followed

Yes No

Review with testing personnel how samples were loaded to rule out misidentification or transposition of samples.

Yes No

Notes: _____

Corrective Action:

It is a Random error

Person Performing Investigation: G. Sateesh Kumar

Date: 16/10/22

Lab Director: Dr. B. Jyothi

Date: 16/10/22

PREPARED & REVIEWED BY:
CONSULTANT PATHOLOGIST: DR. B. JYOTHI

APPROVED & ISSUED BY:
LAB HEAD: DR. B. JYOTHI

B. Jyothi

B. Jyothi

CONTROLLED COPY



TELANGANA DIAGNOSTICS

Form: TD/QSP/08-EQCAR

TITLE

EQAS CORRECTIVE ACTION FORM

Issue No. 01

Page 1 of 1

INVESTIGATION SUMMARY: ROOT CAUSE

Pre-analytic Phase of Testing	Analytic Phase of Testing	Post-Analytic Phase of Testing
<input type="checkbox"/> PROBLEM WITH PT SAMPLE <input type="checkbox"/> SAMPLE PROCESSING <input type="checkbox"/> DATA ENTRY <input type="checkbox"/> OTHER (SPECIFY): _____	<input type="checkbox"/> METHODOLOGICAL PROBLEM <input type="checkbox"/> TECHNICAL PROBLEM <input type="checkbox"/> REAGENT PROBLEM <input type="checkbox"/> CALIBRATOR PROBLEM <input type="checkbox"/> OTHER (SPECIFY): _____	<input type="checkbox"/> CLERICAL ERROR <input type="checkbox"/> REPORTING PROBLEM <input type="checkbox"/> NO EXPLANATION AFTER INVESTIGATION OTHER (SPECIFY): _____

PREVENTION

Preventive action proposed

we will monitor performance of MCHC Parameter closely

Preventive action Plan

we will monitor performance of MCHC Parameter in next cycle of eqas

Responsibility

Date 16/10/22	Testing Personnel Gr. Satish Kumar
Date 16/10/22	Department Technical In charge Dr. B. Jyothi

PREPARED & REVIEWED BY : CONSULTANT PATHLOGIST: DR. B. JYOTHI	APPROVED & ISSUED BY: LAB HEAD: DR. B. JYOTHI
<i>B. Jyothi</i>	<i>B. Jyothi</i>

**TITLE****EQAS CORRECTIVE ACTION FORM**

Issue No. 01

Page 1 of 1

EQAS Details	AIIMS PATHOLOGY
Analyte:	MCV
Month:	AUGUST
Date Sample Tested:	08-09-2022

SPECIMEN HANDLING		
Were specimens received in an acceptable condition?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were specimens stored according to the instructions on the result forms?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the samples hemolyzed?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Were samples tested within the time allowed for sample stability?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
If applicable, were the samples reconstituted correctly?	Yes <input type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
CLERICAL ERRORS		
Were the results transcribed onto the result forms correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results transcribed from the result forms to the website correctly?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded on the correct result form?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Was the correct instrument/reagent/kit selected?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results recorded in the correct units?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Were the results on your evaluation the same as the results you reported?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
QUALITY CONTROL		
Were quality control materials within the acceptable range on the date of PT testing? (Verify the quality control acceptable range in use.)	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
Is there any indication of trending or shifting of the control results?	Yes <input type="checkbox"/>	No <input type="checkbox"/>
Notes: _____		
CALIBRATION		
Were there any problems with the most recent calibration?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
When was the last calibration performed?		
How often is a calibration performed?		
When was the last calibration verification performed?		
Notes: _____		

INSTRUMENT		
Were instrument problems noted the day the samples were tested?	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Has there been any recent maintenance on the analyzer?	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>

PREPARED & REVIEWED BY : CONSULTANT PATHOLOGIST: DR. B. JYOTHI	APPROVED & ISSUED BY: LAB HEAD: DR. B. JYOTHI

CONTROLLED COPY

**TELANGANA DIAGNOSTICS****Form: TD/QSP/08-EQCAR****TITLE****EQAS CORRECTIVE ACTION FORM**

Issue No. 01

Page 1 of 1

Have you contacted your analyzer manufacturer for assistance?

Yes No

Notes: _____

REAGENTS

Were the reagents stored properly?

Yes No

Were the reagents expired or was the open vial stability exceeded?

Yes No

Have there been any changes in reagent manufacturer or formulation?

Yes No

Notes: _____

TESTING PERSONNEL

Date of last competency assessment for testing personnel

Yes No

Review assay procedure and proficiency test sample preparation instructions with testing personnel to ensure that instructions were followed

Yes No

Review with testing personnel how samples were loaded to rule out misidentification or transposition of samples.

Yes No

Notes: _____

Corrective Action:

It is a random error

Person Performing Investigation:

Dr. Sateesh Kumar

Date:

16/10/2022

Lab Director:

Dr. B. Jyothi

Date:

*16/10/2022***PREPARED & REVIEWED BY:
CONSULTANT PATHOLOGIST: DR. B. JYOTHI****APPROVED & ISSUED BY:
LAB HEAD: DR. B. JYOTHI***B. Jyothi**B. Jyothi***CONTROLLED COPY**



TELANGANA DIAGNOSTICS

Form: TD/QSP/08-EQCAR

TITLE

EQAS CORRECTIVE ACTION FORM

Issue No. 01

Page 1 of 1

INVESTIGATION SUMMARY: ROOT CAUSE

Pre-analytic Phase of Testing	Analytic Phase of Testing	Post-Analytic Phase of Testing
<input type="checkbox"/> PROBLEM WITH PT SAMPLE <input type="checkbox"/> SAMPLE PROCESSING <input type="checkbox"/> DATA ENTRY <input type="checkbox"/> OTHER (SPECIFY): _____	<input type="checkbox"/> METHODOLOGICAL PROBLEM <input type="checkbox"/> TECHNICAL PROBLEM <input type="checkbox"/> REAGENT PROBLEM <input type="checkbox"/> CALIBRATOR PROBLEM <input type="checkbox"/> OTHER (SPECIFY): _____	<input type="checkbox"/> CLERICAL ERROR <input type="checkbox"/> REPORTING PROBLEM <input type="checkbox"/> NO EXPLANATION AFTER INVESTIGATION OTHER (SPECIFY): _____

PREVENTION

Preventive action proposed

We will monitor Performance of MCV Parameter closely.

Preventive action Plan

We will monitor performance of MCV Parameter in next cycle of EQAS

Responsibility

Empty box for Responsibility

Date	16/10/2024	Testing Personnel	Dr. Sateesh Kumar
Date		Department Technical In charge	Dr. B. Jyothi

PREPARED & REVIEWED BY: CONSULTANT PATHOLOGIST: DR. B. JYOTHI	APPROVED & ISSUED BY: LAB HEAD: DR. B. JYOTHI
<i>B. Jyothi</i>	<i>B. Jyothi</i>