



L-95, 5th Cross, 1st Main, Kirloskar Colony 3rd Stage,
Water Tank Road, Basaveshwaranagar, Bangalore-560079.

+91 080-2322 3936, 96633 04352

calibration@sarvashree.com

www.sarvashree.com



Cert No. : CC-2291

NABL Accredited Calibration Services as per ISO/IEC 17025:2017

CALIBRATION CERTIFICATE

SS/FF-20/THM-06

Page 1 of 1

1 Name and Address of Customer : M/s. AMRUTHA CLINICAL LABORATORY COMPUTERISED.,
1 st Floor, Shyam Complex, OPP. Shrinidhi Medicals,
Maruthi Veethika, Udupi 576101.

2 Customer Reference

2.1 SRF No : 6450 Dated : 19 July 2022
2.2 Certification No. : SS/22/6450-01 ULR. No : CC229122000009473F
2.3 Date of Calibration : 19 July 2022
2.4 Next Calibration Due : 18 July 2023

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : Hot Air Oven Model : --
3.2 Make : Rotek ID. No. : --
3.3 SI.No : -- Range : Amb to 300 °C
3.4 No. of Pages : 1 LC : 10 °C
3.5 Cal Procedure No. : SOP-THM-06 Location : Lab
3.6 DUC Condition : Satisfactory
3.7 Calibration done at : Onsite

4 Environmental Condition

4.1 Temperature 25 ± 4 °C Humidity 30 to 75 %RH

5 Standards Used for calibration

Sl. No.	Nomenclature	Make / Model	Sl. No	Traceable Cert. No.	Validity
1	4 Wire RTD Sensor With Indicator	Tempsens / Fluke & PT-100 / 8846A	1340 / 3085010	MKS/TH/21-22/311-01	28 February 2023

6 Temperature Calibration Results


Sl. No	Set in °C	DUC Reading in °C	STD Reading in °C	Error Observed in °C	Measurement Uncertainty (±) in °C
1	50	50	49.956	0.044	5.78
2	100	100	99.942	0.058	5.78
3	150	150	149.462	0.538	5.78
4	200	200	199.056	0.944	5.78
5	250	250	248.764	1.236	5.79
6	300	300	298.341	1.659	5.79

7 Note:

- 7.1. Measurement uncertainty reported is at approx 95% confidence level with coverage factor k=2
7.2. Kindly refer Note(s) Section mentioned as below.

Calibrated By

Rakesh J
(Calibration Engineer)



Authorised By

Noushad N
(Lab In-Charge)