

CALIBRATION / TESTING REPORT

CUSTOMER DETAILS: KAMALA SPECIALITY HOSPITAL PRIVATE LIMITED, KOVILPATTI, TAMIL NADU		CERTIFICATE NO	TICS/23/01/002
		EQUIPMENT NAME	MICROSCOPE
		MANUFACTURE MODEL	ADELTA OPTEC
		SERIAL NO	NA
CALIBRATION DATE	09/01/2023	INSTALLED DEPARTMENT	LAB
CALIBRATION DUE DATE	08/01/2024	ASSET NUMBER	KSH-LAB-001
LOCATION OF CALIBRATION	SITE		

- 1. General maintenance required Yes No
- 2. Any Physical damage observed Yes No
- 3. Spare Replacement (if required) Yes No

TEST EQUIPMENT	MANUFACTURE	SERIAL NUMBER	CERTIFICATE NUMBER/VALIDITY
ELECTRICAL SAFTY ANALYZER	DATREND	VPM 18040011	ACCS221441/28-07-2023

*The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

ACCEPTABLE FOR USE CORRECTIVE MAINTENANCE REQUIRED REMOVED FROM USE

Calibrated By : AA
Mandruamorthy

Authorized Sign With Seal

S. Subash
S. SRAIL BABU
TICS
TRANSINDIA CALIBRATION SERVICES PVT LTD

*Electrical safety test result attached

CERTIFICATE NO: TICS/23/01/002


TESTING REPORT			
TEST EQUIPMENT	MANUFACTURE	MODEL	SERIAL NUMBER
ELECTRICAL SAFTY ANALYZER	DATREND	V Pad-mini	VPM 18040011

TEST REPORT –Electrical Safety		Test Standard: IEC 62353		
Parameter	Units	Value	Limits	Status
Main Voltage L1-L2	V	233.4	240±10	PASS
Main Voltage L1-G(PE)	V	234.0	240±10	PASS
Main Voltage L2-G(PE)	V	3.2	<5	PASS
Load Current	A	2.23	As per unit	PASS
Protective Earth Resistance	Ω	0	<0.3	PASS
Point-to-Point Resistance	Ω	0	<0.5	PASS
Equipment Leakage	μA	2.6	Class I - < 500 Class II - < 100	PASS
Applied Part Leakage	μA	0	Type BF - < 5mA Type CF- <50μA*	PASS

* Note: Defibrillator paddles 100 μA, per IEC 60601-2-4

The above mentioned instrument has been tested using standards manufactures recommended protocols, using equipments having traceability to National/International standards.

Mandramonky



PAGE NO : 2/2