



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



PC-1024

Lab Name DR BHARTI PATH LABS

Lab No 16105

Constituent Group Chemistry II

Date of Result Entered : 24/01/2023

PT item Lyophilized Serum

Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	898	169.69	9.93	16.84	176 mg/dL	0.37	1.12
2	UREA II	Urease UV / GLDH II	Transasia / Erba	734	31.57	12.71	4.01	35.59 mg/dL	1.00	0.30
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	570	1.41	11.18	0.16	1.51 mg/dL	0.63	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	799	3.09	15.06	0.46	3.79 mg/dL	1.51	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	719	5.66	9.38	0.53	6.66 g/dL	1.88	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	737	3.30	9.51	0.31	3.89 g/dL	1.88	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	798	4.96	15.45	0.77	4.93 mg/dL	-0.04	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	790	111.91	11.52	12.89	137 mg/dL	1.95	0.92
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	719	97.59	16.60	16.20	80.14 mg/dL	-1.08	1.21

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DR BHARTI PATH LABS
MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
FARIDABAD
HARYANA121004

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name DR BHARTI PATH LABS

Constituent Group Chemistry II

PT item Lyophilized Serum

Lab No 16105

Date of Result Entered : 15/12/2022

Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	856	165.33	10.05	16.62	180.5 mg/dL	0.91	1.14
2	UREA II	Urease UV / GLDH II	Transasia / Erba	735	88.03	12.72	11.20	90.8 mg/dL	0.25	0.83
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	579	1.13	12.84	0.14	1.14 mg/dL	0.07	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	833	2.05	15.55	0.32	2.56 mg/dL	1.60	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	717	4.58	12.91	0.59	4.39 g/dL	-0.32	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	707	2.65	10.97	0.29	2.83 g/dL	0.62	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	771	5.46	11.74	0.64	5.77 mg/dL	0.48	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	813	88.93	14.16	12.60	106.1 mg/dL	1.36	0.88
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	751	79.26	24.12	19.12	71.76 mg/dL	-0.39	1.40

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DR BHARTI PATH LABS
MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
FARIDABAD
HARYANA121004

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2022



Lab Name DR BHARTI PATH LABS

Lab No 16105

Constituent Group Chemistry II

Date of Result Entered : 14/11/2022

PT Item Lyophilized Serum

Date of Report Published : 06/12/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	840	186.15	10.35	19.27	183 mg/dL	-0.16	1.33
2	UREA II	Urease UV / GLDH II	Transasia / Erba	700	51.95	11.66	6.06	56.77 mg/dL	0.80	0.46
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	549	2.62	10.81	0.28	2.49 mg/dL	-0.46	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	756	2.92	13.06	0.38	3.56 mg/dL	1.68	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	724	5.75	8.92	0.51	5.94 g/dL	0.37	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	713	3.30	8.67	0.29	3.53 g/dL	0.80	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	772	5.66	14.64	0.83	5.25 mg/dL	-0.49	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	822	110.52	12.60	13.93	116.6 mg/dL	0.44	0.97
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	760	160.62	13.16	21.14	162.5 mg/dL	0.09	1.53

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DR BHARTI PATH LABS
MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
FARIDABAD
HARYANA121004

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - OCTOBER 2022



PC-1024

Lab Name: DR BHARTI PATH LABS
 Constituent Group: Chemistry II
 PT item: Lyophilized Serum

Lab No: 16105
 Date of Result Entered: 14/10/2022
 Date of Report Published: 03/11/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	854	363.74	10.40	37.83	409 mg/dL	1.20	2.59
2	UREA II	Urease UV / GLDH II	Transasia / Erba	671	34.43	10.58	3.64	41.1 mg/dL	1.83	0.28
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	555	1.49	14.04	0.21	1.52 mg/dL	0.14	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrossik II	Transasia / Erba	714	1.46	19.22	0.28	2.12 mg/dL	2.36	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	666	5.77	7.82	0.45	5.69 g/dL	-0.18	0.03
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	717	3.31	8.64	0.29	3.45 g/dL	0.49	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	685	7.43	10.59	0.79	7.87 mg/dL	0.56	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	790	115.96	10.02	11.62	139.1 mg/dL	1.99	0.83
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	827	210.76	13.41	28.26	226.8 mg/dL	0.57	1.97

PCA

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 DR BHARTI PATH LABS
 MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
 FARIDABAD
 HARYANA121004

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

00 PM



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2022



PC-1024

Lab Name DR BHARTI PATH LABS
 Constituent Group Chemistry II
 PT item Lyophilized Serum

Lab No 16105
 Date of Result Entered : 15/09/2022
 Date of Report Published : 05/10/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	886	188.05	9.72	18.27	181.9 mg/dL	-0.34	1.23
2	UREA II	Urease UV / GLDH II	Transasia / Erba	680	52.25	9.94	5.19	58.92 mg/dL	1.28	0.40
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	562	2.62	10.95	0.29	2.34 mg/dL	-0.98	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	827	2.94	13.86	0.41	3.61 mg/dL	1.64	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	690	5.72	8.35	0.48	5.8 g/dL	0.17	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	714	3.33	9.02	0.30	3.57 g/dL	0.80	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	782	5.67	13.33	0.76	6.34 mg/dL	0.89	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	789	109.84	10.64	11.68	120.3 mg/dL	0.90	0.83
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	764	161.37	11.70	18.88	172.5 mg/dL	0.59	1.37

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 DR BHARTI PATH LABS
 MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
 FARIDABAD
 HARYANA121004

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2022



Lab Name: DR BHARTI PATH LABS Lab No: 16105
 Constituent Group: Chemistry II Date of Result Entered: 16/08/2022
 PT item: Lyophilized Serum Date of Report Published: 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	GLUCOSE II	GOD-POD II	Transasia / Erba	856	109.37	10.01	10.95	133.4 mg/dL	2.20	0.75	✓
2	UREA II	Urease UV / GLDH II	Transasia / Erba	639	23.85	14.71	3.51	37.3 mg/dL	3.83	0.28	✓
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	538	6.31	11.80	0.74	6.26 mg/dL	-0.07	0.06	✓
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	787	4.61	11.51	0.53	4.97 mg/dL	0.68	0.04	✓
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	747	6.06	8.39	0.51	7.01 g/dL	1.87	0.04	✓
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	702	3.42	8.55	0.29	3.91 g/dL	1.67	0.02	✓
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	610	4.08	13.18	0.54	8.45 mg/dL	8.09	0.04	✓
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	731	112.93	10.40	11.74	75.8 mg/dL	-3.16	0.87	✓

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 DR BHARTI PATH LABS
 MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
 FARIDABAD
 HARYANA 121004

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

11:30 AM



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - JULY 2022

Lab Name **DR BHARTI PATH LABS**

Constituent Group **Chemistry II**

PT item **Lyophilized Serum**

Lab No **16105**

Date of Result Entered : **12/07/2022**

Date of Report Published : **05/08/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	841	313.52	11.25	35.28	390 mg/dL	2.17	2.43
2	UREA II	Urease UV / GLDH II	Transasia / Erba	698	59.17	10.93	6.46	85.56 mg/dL	4.09	0.49
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	552	3.38	10.92	0.37	3.42 mg/dL	0.11	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	756	1.82	18.13	0.33	2.29 mg/dL	1.43	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	729	5.98	9.10	0.54	6.07 g/dL	0.17	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	718	3.39	8.38	0.28	4.2 g/dL	2.85	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	772	7.91	12.86	1.02	8.16 mg/dL	0.25	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	826	116.69	10.53	12.28	142 mg/dL	2.06	0.85

ref

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DR BHARTI PATH LABS
MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
FARIDABAD
HARYANA121004

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not subcontract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - JUNE 2022



Lab Name: DR BHARTI PATH LABS
 Constituent Group: Chemistry II
 PT item: Lyophilized Serum

Lab No: 16105
 Date of Result Entered: 16/06/2022
 Date of Report Published: 07/07/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	849	86.39	9.86	8.52	89.8 mg/dL	0.40	0.58
2	UREA II	Urease UV / GLDH II	Transasia / Erba	672	44.59	10.74	4.79	64.4 mg/dL	4.14	0.37
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	554	1.94	11.95	0.23	1.69 mg/dL	-1.08	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	735	3.82	10.81	0.41	3.94 mg/dL	0.29	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	723	6.14	8.88	0.55	6.2 g/dL	0.11	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	700	3.51	8.77	0.31	3.72 g/dL	0.68	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	742	6.84	12.50	0.86	11 mg/dL	4.84	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	810	120.81	12.10	14.62	115 mg/dL	-0.40	1.03
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	736	152.58	12.98	19.81	183.1 mg/dL	1.54	1.46

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 DR BHARTI PATH LABS
 MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
 FARIDABAD
 HARYANA 121004

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL-QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2022



Lab Name DR BHARTI PATH LABS

Lab No 18105

Constituent Group Chemistry II

Date of Result Entered : 07/05/2022

PT item Lyophilized Serum

Date of Report Published : 08/06/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	826	153.23	10.59	16.23	143.1 mg/dL	-0.62	1.13
2	UREA II	Urease UV / GLDH II	Transasia / Erba	628	121.30	12.77	15.49	149.6 mg/dL	-1.83	1.24
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	519	7.94	12.39	0.98	8.69 mg/dL	0.76	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendressik II	Transasia / Erba	669	1.71	18.52	0.32	2.07 mg/dL	1.14	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	672	6.24	8.05	0.50	6.76 g/dL	1.04	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	678	3.50	8.73	0.31	3.4 g/dL	-0.33	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	721	5.78	13.22	0.76	8.41 mg/dL	3.45	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	748	121.76	10.15	12.36	94.36 mg/dL	-2.22	0.90
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	661	89.83	18.41	16.54	117.3 mg/dL	1.66	1.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

DR BHARTI PATH LABS
MOHNA ROAD (IN FRONT OF PUNJAB NATIONAL BANK) BALLABHGARH
FARIDABAD
HARYANA121004

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****