

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony,
Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33ADSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Annai Diagnostics, #Krishnarayar Tank Road, opp. To Harley Rram Hosp. Madurai-01.	Certificate No:	AMS/LAB/289/2835
		Issuing Date:	18/01/2023
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	AN/LAB/CEN/01
Make:	C-Poulo	Department:	Laboratory
Model:	2081	Calibrated On:	17/01/2023
Sl. No:	3335	Calibration Due Suggested:	16/08/2024
Type:	Manual- Digital	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 4000RPM		

Traceability Details:

Sl.No.	Test Equipment**	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/22/ATL/0432-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023
3	Stop Watch *	Nivia / JS 307	PC/01/22/ATL/0134-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (Positions) RPM	Standard Reading (RPM)	Correction (RPM)
1	1000	1020	+20
2	2000	2040	+40
3	2500	2580	+80
4	3000	3060	+60
5	4000	4070	+70

Allowable Tolerance: Set by Customer

Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

Approved by:

(Technical Manager)

