

To Whom It May Concern

For ISO 15189:2012 and ISO 15189:2014 accredited Laboratories — requirements regarding "Calibration & Verification Procedures" [1]

All In vitro Diagnostics Products which are manufactured and distributed by Roche Diagnostics GmbH and for which a Free-Sales-Certificate is issued, are CE-marked.

The In-Vitro-Diagnostics Directive of the European Union [2A.] which is currently switching to IVD Regulation 2017/746/EU (final timeline: May 26, 2022) [2B.] requires for all CE marked products that the manufacturer assures compliance of the products with the requirements of the mentioned directive or regulation. This means that all processes in development and manufacturing of Roche Diagnostics GmbH products are guided by a Quality Management System. Our Quality Management System is in compliance with the requirements from ISO 13485:2016 [3] and 21 CFR Part 820 [4].

The mentioned regulations and standards require that the production systems and measuring devices used are qualified and the manufacturing and test procedures are validated. This status has to be assured by scheduled maintenance and by regular qualification resp. validation reviews and updates.

All physical quantities, calibrators and controls used in Roche Diagnostic systems are fully traceable to certified standards or reference materials. The performance of all In-vitro diagnostics systems of Roche Diagnostics GmbH at the customer site is assured if regular Quality Control measurements, cleaning and maintenance procedures as described in the instructions for use or service documentation are performed. By having controlled internal procedures and by running the tasks required in the respective user documentation, all In-vitro diagnostics systems of Roche Diagnostics GmbH will be performed as specified during their defined lifetime.

Additional calibration or verification procedures are NOT required by the user in order to assure the specified performance of every system of Roche Diagnostics GmbH. Only if a user deviates from these manufacturer's recommendations, the user have to establish site-specific calibration and verification procedures as part of his accreditation process.

- [1] ISO 15189:2012/ ISO 15189:2014 Medical laboratories — Requirements for quality and competence
- [2] A. Directive 98/79/EC of the European Parliament and of the Council of the 27 October 1998 on vitro diagnostics medical devices;  
B. IVD Regulation 2017/746/EU of the European Parliament and of the Council of 5 April 2017 on in vitro diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU
- [3] EN ISO 13485:2016 Medical devices — Quality management systems-Requirements for regulatory purposes
- [4] CFR Part 820, Quality System regulations 21 Regulations on medical devices

Mannheim, 10. August 2021

Sincerely,

Roche Diagnostics GmbH

*i.V./on behalf of the company*

*ppa/on behalf of the company*

  
ECA5294AC4E94AF...

Andrea Weber  
Manager Global Regulatory Affairs  
Centralised and Point of Care Solutions

  
A7F0BA9FE91A46A...

Ralf Zielenski  
Head Q&R Compliance, PRRC RDG  
Centralised and Point of Care Solutions

**Roche Diagnostics GmbH**  
Sandhofer Straße 116  
D-68305 Mannheim



## cobas® pure integrated solutions

Qualification Service  
Installation Qualification / Operation Qualification (v.2.0)



**cobas®**



## cobas® pure integrated solutions

### General Information

Country: INDIA  
Customer Name: Redclieff Lifetech Pvt Ltd  
Customer Address: Saltlake ,Kolkata 700091  
Person Responsible for Quality Assurance: MD KISMAT ANSARI

### System Information

cobas pure	<c 303><e 402>		
	S/N	IP Address	
cobas pure Control unit		172.18.38.	230
SU		172.18.38.	245
303	2145-07	172.18.38.	231
402	2143-02	172.18.38.	232

Host provider:  
cobas IT firewall:  
Control Unit Software Version: 01.02





Qualification Service  
Installation Qualification / Operation Qualification (v 2.0)

Installation Information

Installation Start Date: [Redacted]

First installation:  yes

Reconfiguration: From [Redacted] To [Redacted]

Renovation: From [Redacted] To [Redacted]

Roche Responsible Representative: [Redacted]





**cobas® pure integrated solutions**

## **Installation Qualification**



**cobas®**



## Installation Qualification:

This document forms the basis of the Qualification Services Certificate. It certifies that the instrument is installed according to the manufacturer's specifications. The report presents and documents the test procedures, the documentation, reference and acceptance criteria used to verify that the system is installed according specifications. The report demonstrated that all installation qualification criteria have been met satisfactorily.

**Notice:** The following tests are to be carried out by trained Roche personnel only.

**Purpose:** The purpose of this test is to confirm that the instrument was delivered undamaged and installed correctly.

Test #	Test	Pass Fail	Signature Date
IQ.1.1	User assistance available	Pass	
IQ.1.2	Environmental parameters met	Pass	
IQ.1.3	Instrument delivered undamaged and complete	Pass	
IQ.1.4	Transport locking successfully removed	Pass	
IQ.1.5	All connections correctly installed	Pass	
IQ.1.6	Instrument positioned according to Installation Manual	Pass	
IQ.1.7	Instrument boot process successfully	Pass	
IQ.1.8	Checksum according to specification	Pass	
IQ.1.9	Mechanical adjustments complete	Pass	
IQ.1.10	Auxiliary components positioned	Pass	
IQ.1.11	Instrument installation check	Pass	
IQ.1.12	Host communication settings checked	yes	

Test #	Test	Pass Fail	Signature Date
IQ.2	Installation Qualification for cobas <ISE>	yes	
IQ.3	Installation Qualification for cobas <c 303>	yes	
IQ.4	Installation Qualification for cobas <e 402>	yes	
IQ.5	Installation Qualification for cobas link	yes	

**cobas**<sup>®</sup>



**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

<b>Deviation #1</b>
<b>Investigation</b>
<b>Action taken</b>
<b>Deviation resolved satisfactorily?</b> specify

<b>Deviation #2</b>
<b>Investigation</b>
<b>Action taken</b>
<b>Deviation resolved satisfactorily?</b> specify

<b>Deviation #3</b>
<b>Investigation</b>
<b>Action taken</b>
<b>Deviation resolved satisfactorily?</b> specify







## Operational Qualification:

This document is the basis of the Qualification Service Certificate. It certifies that the instrument is operating according to the manufacture's specifications. This report presents and documents the test procedures, documentation, references and acceptance criteria used to verify that the specified system is operating according the specifications. The report demonstrates that all operational qualification criteria have been met satisfactorily.

**Notice:** The following tests are to be carried out by trained Roche personnel only.

**Purpose:** The purpose of this test is to check that the modules are operating in accordance with the

Test #	Test	Pass Fail	Signature Date
OQ.1	Calibration successfully	Pass	
OQ.2	Quality Control successfully	Pass	
OQ.3	Accuracy check successfully	Pass	

**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

**Deviation #1**

Investigation

Action taken

Deviation resolved satisfactorily? specify

**Deviation #2**

Investigation
Action taken
Deviation resolved satisfactorily? <span style="float: right;">specify</span>



Qualification Service  
Installation Qualification / Operation Qualification (v.2.0)

Page 6 of 6

### Conclusion

- All test results are acceptable. yes
- Any deviation or non-conformances observed have been recorded as a deviation and the relevant forms completed. yes
- All acceptance criteria have been met. This equipment is deemed acceptable and the unit is approved for its intended use. yes

### Comments

---

---

---

---

---

---

---

Completed by Roche Representative Date 13/6/22  
 Print Name SASWATA DASGUPTA Signature [Signature]

Reviewed by Customer Contact Date 14/06/22  
 Print Name SUVANKAR BHATTACHARJEE Signature [Signature]

Reviewed by Customer Quality Assurance Date 14/06/22  
 Print Name MD KISMAT ANSARI Signature [Signature]





## Installation Qualification for cobas® pure <ISE>

### Description

IQ.2.2	Mechanical adjustments complete	
	All mechanical adjustments for ISE mechanical parts are carried out	Pass
	Adjustment check during mechanical check function	Pass
IQ.2.3	Auxiliary components positioned	
	ISE Reagents are loaded	yes
IQ.2.4	Gear pump adjustment	
	Gear pump adjustment executed (attached printout)	yes
IQ.2.5	Instrument installation check	
	ISE Check 20 times (attached printout)	yes
IQ.2.6	Application installation	
	Download of applications from cobas link: Na (29070) K (29080) CI (29090)	Pass

**cobas**®



**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

<b>Deviation #1</b>
<b>Investigation</b>
<b>Action taken</b>
<b>Deviation resolved satisfactorily?</b> specify

<b>Deviation #2</b>
<b>Investigation</b>
<b>Action taken</b>
<b>Deviation resolved satisfactorily?</b> specify





## Installation Qualification for cobas® pure <c 303>:

### Description

<b>IQ.3.1</b>	<b>Function check of c 303 module according to specifications</b>	
	System layout is according to the description in the manual	Pass
	c 303 AU is installed according to the installation manual with official tools	Pass
<b>IQ.3.2</b>	<b>Mechanical adjustments complete</b>	
	All mechanical adjustments for the different c 303 mechanical parts are carried out	Pass
<b>IQ.3.3</b>	<b>Probes and consumables installation</b>	
	Sample Probe and Reagent probe are installed	Pass
	Reaction cells are installed	Pass
<b>IQ.3.4</b>	<b>Instrument installation check</b>	
	Basic and Acid wash bottles are loaded	Pass
	Water pressure: Main pump 50.0–60.0 kPa, Gear Pump 320 kpa	Pass
	Load ECO-D c pack green	Pass
	Exchange incubation bath water	Pass
	Water flow of rinse stations as well as consumption of the detergents were adjusted	Pass
	Air purge for syringes and reagents	Pass
	Photometer check (attached printout)	Pass
	Cell Blank Measurement (attached printout)	Pass
	Incubation water bath temperature 37 °C ± 0.1 °C	Pass

cobas®





	Adjustment check during mechanical check function	Pass
<b>IQ 3.5</b>	<b>Gear pump adjustment</b>	
	Gear pump adjustment (attached printout)	yes
<b>IQ 3.6</b>	<b>Application Installation</b>	
	Download Special Wash (all)	Pass
	Download of Auxiliary Reagents from cobas link ECO-D, NAOHD, SMS and PYP (PYP if ASTP2 is not available) Menu>System>Auxiliary Reagent Packs>Download	Pass
	Download of applications from cobas link ASTP (20220) (if ASTP2 (20230) is not available), CHOL2 (20411) CREJ2 (20470) GLUC3 (20630) TP2 (21110) CONA-P2 (20993) CONA-R1 (21280) INST-S1 (21290) INST-R1 (21291) Menu>Application>Download	Pass
<b>IQ 3.7</b>	<b>Instrument check</b>	
	Instrument Check (attached printout)	Pass
<b>IQ 3.8</b>	<b>Backup of adjustment data</b>	
	Adjustments data backed up	Pass

## Installation Qualification for cobas® pure <e 402>:

### Description

<b>IQ.4.1</b>	<b>Function check of e 402 module according to specifications</b>	
	e 402 AU is installed according to the installation manual with official tools	Pass
<b>IQ.4.2</b>	<b>Mechanical adjustments complete</b>	
	All mechanical adjustments for the different e 402 mechanical parts are carried out	Pass
	Adjustment check during mechanical check function	Pass
	Water pressure: Gear pump 320 kPa, Main pump 50.0–60.0 kPa	Pass
	Water flow of all rinse stations and wash station was adjusted and validated	Pass
<b>IQ.4.3</b>	<b>Auxiliary components installed</b>	
	Sample probe, reagent probe, microbeads mixer, measuring cell, sipper probe and pre-wash sipper probe installed	yes
	Waste liner, CC/PC cups, CleanCell, ProCell, PreClean and Assay Cup&Tip trays loaded	yes
	System prime and system air purge for syringes and reagents	Pass
<b>IQ.4.4</b>	<b>Instrument installation check</b>	
	Temperatures within specifications	Pass
	No alarms during check	Pass
	Air Aspiration Calibration	Pass
	PMT Setting	Pass
	Blank Cell calibration (attached printout)	Pass
	Instrument Check (attached printout)	Pass
<b>IQ.4.5</b>	<b>Application installation</b>	
	"Elecsys TSH for Instrument Check" * e-pack does not need TSH application to be downloaded. * GMMI : 0702 8091 200	Pass
	- For Precision Check (OQ.3.3) you need a New TSH e-pack and download TSH (10172) application from cobas link. IC TSH can not be used for Precision Check.	
<b>IQ.4.6</b>	<b>Backup of adjustment data</b>	
	Adjustments data backed up	Pass

**cobas**®



**Deviation Report:** Any discrepancies found during the installation must be documented in the space below.

**Deviation #1**

**Investigation**

**Action taken**

**Deviation resolved satisfactorily?** specify

**Deviation #2**

**Investigation**

**Action taken**

**Deviation resolved satisfactorily?** specify

**Deviation #3**

**Investigation**

**Action taken**

**Deviation resolved satisfactorily?** specify







## Installation Qualification for cobasLink

### Description

<b>IQ.5.1</b>	<b>cobasLink connectivity test</b>	
	cobasLink is installed according to the cobas Link cobas pure manual	YES
	Internet connection is available	YES
<b>IQ.5.2</b>	<b>cobasLink configuration</b>	
	cobas link Configurator/System Check Latest patches installed	OK
	General settings are entered according to cobasLink Manual (Laboratory/cobasLink/Utilities incl. certificate )	OK
	cobas link Configurator/System Check (green traffic light)	OK
	The CU - configuration was sent to CL and is visible @ cobas Link configurator/Query Tool/RSI2 > Execute	YES
<b>IQ.5.3</b>	<b>cobasLink Initiate Upload/ Download</b>	
	The configuration was sent to the TSN server with "Initiate Upload"	YES
	The application files arrive after "Initiate Download"	YES
	Perform Sync with cobas link Maintenance>Service>Sync with cobas link Wait 30min	YES

**cobas**<sup>®</sup>





**Deviation Report:** Any discrepancies found during the installation must be documented in the space below. Roche personnel will then investigate the deviation and decide upon the most appropriate action to be taken.

**Deviation #1**

**Investigation**

**Action taken**

Deviation resolved satisfactorily? specify

**Deviation #2**

**Investigation**

**Action taken**

Deviation resolved satisfactorily? specify

**cobas**<sup>®</sup>



**cobas® pure integrated solutions**  
**Operational Qualification**



**cobas®**

## Operational Qualification:

**Notice:** The steps described in OQ.1 have to be carried out after a new system installation and after any repair action which requires additional calibration.

If the service action does not affect the measurement performance, only apply steps OQ.2 and OQ.3 of the Operation Qualification.

### Description

#### OQ 1 Calibration

Calibration of all photometric parameters successful (attached printout) yes

Calibration of ISE parameters successful (attached printout) yes

Calibration of all Immuno parameters successful (attached printout) yes

Specify the type of calibrator used:

C.F.A.S. lot : 539942

TSH CalSet lot : 575647

#### OQ.2 Quality Control

Specify the type of control used:

PCCC1 lot : 525027

PCCC2 lot : 519198

PCU1 lot : 556403

PCU2 lot : 556405

QC of all photometric parameters within acceptable range (attached printout) yes

QC of ISE parameters within acceptable range (attached printout) yes

QC of Immuno parameters within acceptable range (attached printout) yes

#### OQ.3.1 Precision check for ISE

Perform precision check using PCCC1 n=21

	Number of det.	Expected < CV	Actual CV
Na	21	1.00%	0.29%
K	21	1.20%	0.30%
Cl	21	1.70%	0.25%

Expected precision CV values are only to judge performance of newly installed analyzer, for official specification please refer to assay specific Method sheet.

Precision check for ISE was within acceptable range yes



**OQ.3.2 Precision check for Photometric Assays**

Perform precision check using PCCC1 n=21

	Number of det.	Expected < CV	Actual CV
ASTP/ASTP2	21	2.00%	1.99%
GLUC3	21	1.00%	0.49%
CREJ2	21	2.50%	1.64%
TP2	21	1.00%	0.29%
CHOL2	21	1.00%	0.65%

Expected precision CV values are only to judge performance of newly installed analyzer, for official specification please refer to assay specific Method sheet.

Precision check for Photometric Assays was within acceptable range

yes

**OQ.3.3 Precision check for Immunology Assays**

Perform precision check using PCU1 n=21 per Channel

New TSH (10172) Reagent	Number of det.	Expected < CV	Actual CV
TSH Ch.1	21	5.00%	0.92%

Precision

yes

**OQ 3.4 Trace Doc**

Calibration

yes

QC results

yes

Sample

yes



Deviation #1
Investigation
Action taken
Deviation resolved satisfactorily? <small>specify</small>



Page 1  
Attachments



cobas®





**cobas® pure integrated solutions**  
**Attachments**



**cobas®**





Sample Results Report

Result Type: First

Sorted by: Date

Date

Sequence No: RadclID Pos

SampleID / QC Lot

User ID

Registered

23/02/2023

10:27:32

Comment: Q002130 Q30001-1

Name / Lot:

PCCC1 / 525027

ADMIN

Test Unit	Sample type	Result Data	Alarm	A / U	Priority	R.P. INO	R.P. Serial NO	Protocol: Ser / Elfdrtg
ALB2-G	g/dL	3.28		CC	Current	R1-R3	608713	0000584
ALP2	U/L	90.9		CC	Current	R1-R3	668539	0002241
ALTP	U/L	46.3		CC	Current	R1-R3	661052	0014638
						R2	654893	0002984
AMWL2	U/L	84.1		CC	Current	R1-R3	630349	0002956
APOAT	mg/dL	102	OBS.RR	CC	Current	R1-R3	588324	0000434
APOBT	mg/dL	41.7	OBS.RR	CC	Current	R1-R3	631378	0000957

*[Handwritten Signature]*  
23/02/2023

System: cobas pure

Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No: Q002130 Rack ID: Pos: Q30001-1 Sample ID / QC Lot: PCCCI / 525027  
 Comment: Name / Lot: ADMIN Registered: 23/02/2023 10:27:32

Test Unit	Sample Type	Result Data Alarm	Dilution	Unit	Priority	R1-R3	R1-R3	R1-R3	R1-R3	Registered
ASTP U/L	Ser/Pl	45.9		CC	Current	R1-R3 R2	663695 654893	0015390 0002984		
ASTP U/L	Ser/Pl	46.7		CC	SBI	R1-R3 R2	663695 654893	0017958 0002984		
BILD2-D mg/dL	Ser/Pl	0.703		CC	Current	R1-R2	652625	0001519		
BILT3 mg/dL	Ser/Pl	0.864		CC	Current	R1-R3	622281	0004606		
CA2 mg/dL	Ser/Pl		Samp:5	CC	Current					
CHOL2-I mg/dL	Ser/Pl	91.7		CC	Current	R1	640736	0002978		

*Handwritten signature*

System: cobas pure Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No: Q002130 Rack ID: Q30001-1 Sample ID: PCC1 / 525027  
 Comment: Name / Lot: PCC1 / 525027  
 Registered: 23/02/2023 10:27:32  
 User ID: ADMIN

Test	Unit	Sample Type	Result	Data	Alarm	Dilution	A.U.	Priority	Ref. Lot	R1-R3	Serial No.	Protocol	Unit / Dilution	Registered
CREJ2	mg/dL	Ser/Pl	1.08	CC	Current	R1-R3	627775	0003592						
CRP4	mg/L	Ser/Pl	9.08	CC	Current	R1-R3	645296	0004112						
CRP-HS	mg/L	Ser/Pl	10.4	CC	Current	R1-R3	626644	0000015						
GGT2-I	U/L	Ser/Pl	55.2	CC	Current	R1-R3	653840	0006114						
GLUC3	mg/dL	Ser/Pl	101	CC	Current	R1-R3	629487	0003425						
HDLC4	mg/dL	Ser/Pl	29.5	CC	Current	R1-R3	626113	0006797						
IRON2	µg/dL	Ser/Pl	105	CC	Current	R1-R3	641946	0000593						

System: cobas pure

Serial No.: 2154-03

*Signature*

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No.	Rack ID	Pos	Sample ID / QC Lot	User ID	Registered
Q002130	Q30001-1		Name / Lot: PCCC1 / 525027	ADMIN	23/02/2023 10:27:32

Test	Unit	Sample Type	Result Data/Alarm	Alt	Priority	R.P. Lot	R.P. Serial No.	Protocol Lot / Electrode
ISE NA-S	mmol/L	Serum	105.8 QC Err					
				ISE	Current	IS	624115	0000700
						DIL	650042	0010909
						REF	661046	0008910
								E2196 X6546

LDHI2	U/L	Ser/Pl	Samp.S	CC	Current	R1-R3	572897	0001121
LDLC3	mg/dL	Ser/Pl	57.0	CC	Current	R1-R3	572897	0001121
LIP	U/L	Ser/Pl	Samp.S	CC	Current	R1-R3	647416	0000052
PHOS2	mg/dL	Ser/Pl	3.92	CC	Current	R1-R3	653804	0004108
TP2	g/dL	Ser/Pl	4.68	CC	Current	R1-R3	635652	0005917



System: cobas pure

Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No: Q002130 Rack ID: Q30001-1 Pos: PCC1 / 525027 Name / Lot: PCC1 / 525027  
 Comment: User ID: ADMIN Registered: 23/02/2023 10:27:32

Test Unit	Sample Type	Result Data	Alarm	A.U.	Priority	R.P. Lot	R.P. Serial No.	Reg Call No. / Alternate
TRIGL	Ser/Pl	109		CC	Current	R1	643364	0003033
UA2	Ser/Pl	4.75		CC	Current	R1-R3	627564	0002726
UIBC-1	Ser/Pl	206		CC	Current	R1-R3	637859	0015422
UIBC-1	Ser/Pl	201		CC	SB1	R1-R3	637859	0014774
UREAL	Ser/Pl	37.6		CC	Current	R1-R3	651528	0008399
UREAL	Ser/Pl	39.7		CC	SB1	R1-R3	651528	0008658

*[Handwritten Signature]*  
23/02/2023

Symbols attached to "Test"

\*: Rerun test

~: eFlow subresult supplemented test for formula calculated test and Hb/HbA1c test

System: cobas pure

Serial No.: 2154-03



Sample Results Report

Result Type: First Sorted by: Date

Sequence No: Rad/ID Pos Sample ID/Order ID Registered: 23/02/2023 13:23:08

Comment: Q002131 Q30001-1 Name / Lot: PCCC1 / 525027 ADMIN

Test	Unit	Sample Type	Result	Date/Alarm	Dilution	A	U	Priority	R/P Lot	R/P Serial No.	Physical Lot / Electrode
------	------	-------------	--------	------------	----------	---	---	----------	---------	----------------	--------------------------

ISE CL	mmol/L	Ser/Pl	89.1	ISE	Current	IS			624115	0000700	X2918 X6546
						DIL			650042	0010909	
						REF			661046	0008910	

ISE K	mmol/L	Ser/Pl	3.72	ISE	Current	IS			624115	0000700	M6811 X6546
						DIL			650042	0010909	
						REF			661046	0008910	


ISE NA	mmol/L	Ser/Pl	117.2	ISE	Current	IS			624115	0000700	E2196 X6546
						DIL			650042	0010909	
						REF			661046	0008910	

Symbols attached to "Test"

\*: Rerun test

~: eFlow subresult, supplemented test for formula calculated test and Hb/HbA1c test

System: cobas pure Serial No.: 2154-03



Sample Results Report

Result Type: First Sorted by: Date Registered: 23/02/2023 13:35:42

Sequence No: Rad/Id Pos Sample ID: Q30001-1 Name / Lot: PCCCL1 / 525027 User ID: ADMIN

Comment: Q002132

Test Unit	Sample type	Result Data Alarm	Priority	R1-R3	R1P-Serial No	Protocol ID / Electrode
CA2		9.15	CC	R1-R3	645917	0000592
LDH12		180	CC	R1-R3	628352	0002361
LIP		45.3	CC	R1-R3	647416	0000052

mg/dL	Ser/Pl	U/L	Ser/Pl	U/L	Ser/Pl

*Handwritten signature and date: 23/02/2023*

Symbols attached to "Test"

\*: Rerun test

+: eFlow subresult, supplemented test for formula calculated test and Hb/HbA1c test

System: cobas pure

Serial No.: 2154-03



Sample Results Report

Result Type: First Sorted by: Date

Sequence No: Rad/ID: Pos Sample ID: Q00130

Comment: Q001130 Q30001-2 Name / Lot: PCCC2 / 519198

Registered: 23/02/2023 10:27:32

ADMIN

Test Unit	Sample Type	Result Data Alarm	Alt	Priority	Req. No.	R.P. Serial No.	Original Lot / Expiry Date
ALB2-G	g/dL	Ser/Pl	5.10	CC	Current	R1-R3	608713 0000584
ALP2	U/L	Ser/Pl	240	CC	Current	R1-R3	668539 0002241
ALTP	U/L	Ser/Pl	123	CC	Current	R1-R3	661052 0014638
						R2	654893 0002984
AMYL2	U/L	Ser/Pl	186	CC	Current	R1-R3	630349 0002956
APOAT	mg/dL	Ser/Pl	160 OBS.RR	CC	Current	R1-R3	588324 0000434
APOBT	mg/dL	Ser/Pl	74.4 OBS.RR	CC	Current	R1-R3	631378 0000957

*Handwritten signature*

System: cobas pure

Serial No.: 2154-03



# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No: Req'd ID Pos Sample ID / QC Lot User ID Registered

Comment: Q001130 Q30001-2 Name / Lot: PCCC2 / 519198 ADMIN 23/02/2023 10:27:32

Test Unit	Sample Type	Dilution	A.U.	Priority	R1	R2	R3	R1-R2	R1-R3	R1-R3	R.P. Serial No.	Prog Call Lot / Electrode
ASTP	U/L		149	CC	Current	R1-R3	663695	0015390	654893	0002984		
ASTP	U/L		149	CC	SB1	R1-R3	663695	0017958	654893	0002984		
BILD2-D	mg/dL		2.12	CC	Current	R1-R2	652625	0001519				
BILT3	mg/dL		3.46	CC	Current	R1-R3	622281	0004606				
CA2	mg/dL	Samp.S		CC	Current							
CHOL2-I	mg/dL		174	CC	Current	R1	640736	0002978				



System: cobas pure

Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No: Rack ID: Pos: Sample ID / QC Lot: [User ID]

Comment: Name / Lot: PCCC2 / 519198 Registered: 23/02/2023 10:27:32

Q001130 Q30001-2 ADMIN

Test	Unit	Sample Type	Dilution	Result	Alarm	CC	Current	R1-R3	R.P. Lot	R.P. Serial No.	Proced Lot / Reagent
CREJ2	mg/dL	Ser/Pl		4.26		CC	Current	R1-R3	627775	0003592	
CRP4	mg/L	Ser/Pl		50.4		CC	Current	R1-R3	645296	0004112	
GGT2-1	U/L	Ser/Pl		226		CC	Current	R1-R3	653840	0006114	
GLUC3	mg/dL	Ser/Pl		241		CC	Current	R1-R3	629487	0003425	
HDLC4	mg/dL	Ser/Pl		57.3		CC	Current	R1-R3	626113	0006797	
IRON2	µg/dL	Ser/Pl	Samp.S			CC	Current				

*[Handwritten Signature]*

System: cobas pure

Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No	ReckID	Pos	Sample ID/ QC Lot	User ID	Registered
Q001130	Q30001-2		Name / Lot: PCCC2 / 519198	ADMIN	23/02/2023 10:27:32

Test	Unit	Sample Type	Dilution	A.U	Priority	R.P. Lot	R.P. Serial No.	ProCall Lot / Electrode
Result Message								

ISE CL mmo/L Ser/Pl 98.5 ISE Current IS 624115 0000700 X2918 X6546  
 ISE CL-P mmo/L Plasma 99.0 ISE Current IS 624115 0000700 X2918 X6546  
 ISE CL-S mmo/L Serum 99.0 ISE Current IS 624115 0000700 X2918 X6546  
 ISE K mmo/L Ser/Pl 6.67 ISE Current IS 624115 0000700 M6811 X6546

REF 661046 0008910  
 DIL 650042 0010909  
 REF 661046 0008910

System: cobas pure

Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No	Back ID - Pos	Sample ID / Operator	User ID	Registered
Q001130	Q30001-2	Name / Lot:	PCCC2 / 519198	ADMIN
				23/02/2023 10:27:32

Test Unit	Sample Type	Dilution	A.U	Priority	R.P. Lot	R.P. Serial No.	Prog Call Lot / Electrode
ISE K-P	Plasma						
6.68							
ISE K-S	Serum						
6.70							
ISE NA	Ser/Pl						
128.4							
ISE NA-P	Plasma						
128.5							

ISE	Current	IS	DIL	REF	624115	650042	661046	0000700	0010909	0008910	M6811 X6546
ISE K-P	Current	IS	DIL	REF	624115	650042	661046	0000700	0010909	0008910	M6811 X6546
ISE K-S	Current	IS	DIL	REF	624115	650042	661046	0000700	0010909	0008910	M6811 X6546
ISE NA	Current	IS	DIL	REF	624115	650042	661046	0000700	0010909	0008910	M6811 X6546
ISE NA-P	Current	IS	DIL	REF	624115	650042	661046	0000700	0010909	0008910	M6811 X6546

System: cobas pure

Serial No.: 2154-03

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No	Rad ID	Pos	Sample ID / Q.C. Lot	Registered
Q001130	Q30001-2		Name / Lot: PCCC2 / 519198	23/02/2023 10:27:32
				ADMIN

Test	Unit	Sample Type	Dilution	Result Data Alarm	A.U.	Priority	R1.P. Lot	R1.P. Serial No.	Protocol / Lot / Electrode	
Result Message										
ISE NA-S		Serum		129.1	ISE	Current	IS	624115	0000700	E2196 X6546
	mmol/L						DIL	650042	0010909	
							REF	661046	0008910	

Test	Unit	Sample Type	Dilution	Result Data Alarm	A.U.	Priority	R1.P. Lot	R1.P. Serial No.	Protocol / Lot / Electrode
LDHI2	U/L	Ser/Pl		Samps.S	CC	Current			
LDIC3	mg/dL	Ser/Pl		106	CC	Current	R1-R3	572897	0001121
LIP	U/L	Ser/Pl		Samps.S	CC	Current			
PHOS2	mg/dL	Ser/Pl		7.79	CC	Current	R1-R3	653804	0004108
TP2	g/dL	Ser/Pl		7.73	CC	Current	R1-R3	635652	0005917

System: cobas pure

Serial No.: 2154-03

*Handwritten signature and date: 23/02/2023*

# Sample Results Report

User ID: ADMIN

23/02/2023 15:39

Result Type: First Sorted by: Date

Sequence No	Redt/ID	Pos	Sample ID / QC/Lot	Name / Lot	PCC2 / 519198	User ID	Registered
Q001130	Q30001-2					ADMIN	23/02/2023 10:27:32

Test	Unit	Sample Type	Dilution	Alt	Priority	R1/P/Lot	R1/P/Serial No.	Prog/Call Lot / Electrode
TRIGL	mg/dL	Ser/Pl		CC	Current	R1	643364	0003033
UA2	mg/dL	Ser/Pl		CC	Current	R1-R3	627564	0002726
UIBC-1	µg/dL	Ser/Pl		CC	Current	R1-R3	637859	0015422
UIBC-1	µg/dL	Ser/Pl	Samp.S	CC	SBI	R1-R3	637859	0014774
UREAL	mg/dL	Ser/Pl		CC	Current	R1-R3	651528	0008399
UREAL	mg/dL	Ser/Pl		CC	SBI	R1-R3	651528	0008658

Symbols attached to "Test"

\*: eFlow subresult, supplemented test for formula calculated test and Hb/HbA1c test

System: cobas pure Serial No.: 2154-03

*Handwritten signature and date: 23/02/2023*



Sample Results Report

Result Type: First Sorted by: Date

Sequence No. Back ID Pos Sample ID / QC Lot User ID Registered

Comment Q0001-2 Name / Lot: PCCCC2 / 519198 ADMIN 23/02/2023 13:35:42

Test	Unit	Sample Type	Result Data Alarm	Dilution	A.U.	Priority	R.I.P. Lot	R.I.P. Serial No.	Proced Lot / Electrode
CA2	mg/dL	Ser/Pl		13.7		CC	Current	R1-R3	645917 0000592
IRON2	µg/dL	Ser/Pl		233		CC	Current	R1-R3	641946 0000593
LDHI2	U/L	Ser/Pl		322		CC	Current	R1-R3	628352 0002361
LIP	U/L	Ser/Pl		101		CC	Current	R1-R3	647416 0000052

Symbols attached to "Test"

\*: Retrun test

\*\*: eFlow subresult, supplemented test for formula calculated test and Hb/HbA1c test

System: cobas pure

Serial No.: 2154-03

*[Handwritten signature]*  
23/02/2023



Sample Results Report

Result Type: First Sorted by: Date Registered

Sequence No. Rack ID: Pos. Sample ID / Operator User ID: 23/02/2023 13:23:08

Comment: Q001131 Q30001-2 Name / Lot: PCCC2 / 519198 ADMIN

Test	Unit	Sample Type	Result	Date/Alarm	Dilution	AVU	Priority	R.P. Lot	R.P. Serial No.	Protocol Lot / Electrode
ISE CL	mmol/L	Ser/PI	105.6					624115	0000700	X2918 X6546
								650042	0010909	
								661046	0008910	
ISE K	mmol/L	Ser/PI	7.17					624115	0000700	M6811 X6546
								650042	0010909	
								661046	0008910	
ISE NA	mmol/L	Ser/PI	138.4					624115	0000700	E2196 X6546
								650042	0010909	
								661046	0008910	

*Handwritten signature and date: 23/2/2023*

Symbols attached to "Test"

\*: Rerun test

\*\* : eFlow subresult, supplemented test for formula calculated test and Hb/HbA1c test

System: cobas pure

Serial No.: 2154-03