

# Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR

DL No. MDU/4205/20B,MDU/3971/21B

CIN : U33110TN2009PTC072140

## CALIBRATION CERTIFICATE

Customer Name & Address:	M/S, Quality Care Diagnostic Centre, 91, a-1, K.S. Building Trichy Road, Namakkal-637001	Certificate No:	AMS/LAB/289/1954
		Issuing Date:	03/09/2022
		Page No:	1 of 1

### Details of Device under Calibration (DUC):

DUC:	Centrifuge	Id. No:	QCDC/LAB/CEN/001
Make:	ULTIMATE	Department:	Laboratory
Model:	DIGITAL	Calibrated On:	01/09/2022
Sl. No:	----	Calibration Due Suggested:	31/08/2023
Type:	Digital – Manual Control	DUC Condition on Receipt:	OK
Range & Least Count:	0 – 4000 / LC – 1; Timer 10 to 60Mins. In 10Mins. interval		

### Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Digital Tachometer	Mextech / DT-2234C	PC/01/22/ATL/0432-01	30/01/2023
2	Thermo Hygrometer	Beetech/088	PC/01/22/ATL/0265-01	30/01/2023
3	Stop Watch	Nivia / JS 307	PC/01/22/ATL/0134-01	30/01/2023

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

### Test Results: Rotation Per Minute

SL.No.	DUC Reading (RPM)	Standard Reading (RPM)	Correction (RPM)
1	1000	1000	0
2	2000	2000	0
3	2500	2500	0
4	3000	3000	0
5	4000	4000	0

Allowable Tolerance: Set by Customer



Note:

1. Calibration Points were selected as per customer request.
2. Each reading was taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by:

(Technical Manager)