

HORIBA

Explore the future

HORIBA India Private Ltd.

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2023-2024/100737

10th Feb 2023


CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **ABX YUMIZEN H550** bearing serial number:
110YAXH03533 installed at **P.H. Medical Centre, Santacruz** calibrated on **10th Feb 2023**

Calibrator : ABX MINOCAL
Lot No. : CX480
Expiry Date : 05th April 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 09th Feb 2024.



Shrish Dixit

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**