



Abbott Healthcare Pvt. Ltd.
Floor 13-19, Godrej BKC, Plot C-68
G Block, Bandra Kurla Complex
Near MCA Club, Bandra East
Mumbai - 400051. India

Registered Office:
4, Corporate Park,
Sion Trombay Road,
Mumbai - 400071. India

T : +91 22 3816 2000 / 1010
F : +91 22 3816 2400 / 2401
Email : webmaster@abbott.co.in
Site : www.abbottdiagnostics.com
CIN : U24200MH1997PTC104834

Calibration Certificate

25th Nov 2022

To,
PH Medical Center
Biochem Department
Mumbai

Dear Sir/Madam,

This is with reference to the Alinity Ci Analyzer (Sr. No. AC03853 & AI21676) installed in your laboratory.

Please find below the details of the instrument Calibrations and the Verifications performed.

The next calibration will be due on 24th November 2023

Instrument : Alinity Ci Immunochemistry analyzer.
Serial number : AC03853 & AI21676
Calibration performed on : 25th Nov 2022

1. System voltage checks performed using Digital Multimeter and confirmed to specifications.
2. Sample & Reagent Pipettor Calibration Passed.
3. Reagent Transport Calibration Passed.
4. Carrier Transport Calibration Passed
5. Wash Zone Calibration Passed

Thanking you,

Yours Truly,

Arun Kuriakose
Customer Support Specialist