

STATE REFERENCE LABORATORY (SRL),SVNGMC ,YAVATMAL
Timer Calibration Verification Record Sheet

Location: ICTC Lab, RH Wani

Property No: DT 109

Procedure using Timer: Inhouse verification of timer using certified standard calibrated Timer

Date: 25/07/2022

Due Date:24/07/2023

Certified timer ID: SRL/MICRO/SVNGMC/37

Verification Done by: Ms Urmila R. Chambhare, Technical Officer, SRL,SVNGMC,Yavatmal


Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference
							900	900	0
1	300	300	0	600	599	-1	900	900	0
2	300	299	-1	600	601	1	900	899	-1
3	300	300	0	600	601	1	900	900	0
4	300	300	0	600	601	1	900	900	0
5	300	300	0	600	600	0	900	900	0
6	300	300	0	600	601	1	900	900	0
7	300	299	-1	600	600	0	900	900	0
8	300	300	0	600	600	0	900	899	-1
9	300	299	-1	600	600	0	900	901	1
10	300	299	-1	600	600	0	900	899.9	-0.1
Mean	300	299.6	-0.4	600	600.3	0.3	900		
SD	-	0.52	-	-	0.67	-	-	0.57	-
%CV	-	0.172362	-	-	0.112435	-	-	0.063079	-
% Accuracy	-	-0.133333	-	-	0.05	-	-	-0.011111	-

Accept/ Reject: Accepted since variation (%CV) is within 2% of all three times set

Final Interpretation: Timer is in good working condition

Action Taken: -

Remark: -


Urmila R. Chambhare
 Technical Officer
 SRL (Microbiology)
 Shri V.N.G.M.C, Yavatmal