



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JANUARY 2023



Lab Name	SYMBION VIP DIAGNOSTICS	Lab No	1803
Constituent Group	HbA1c	Date of Result Entered :	06/01/2023
PT item	Lyophilized Whole blood	Date of Report Published :	09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	HbA1c	HPLC-Ion Exchange	Biorad (D10 / D100 / Variant Turbo)	620	5.50	3.22	0.18	5.3 %	-1.13	0.01

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

SYMBION VIP DIAGNOSTICS

1-SPRING FIELD BUNGLOW, OPP. AKSHARDHAM COMPLEX, JUDGES BUNGLOW ROAD, BODAK DEV

AHEMDABAD

GUJARAT380054

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****