



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 534

Distribution No.: 157-B

Month/Year: July/2022

Instrument ID: DxH-800 BECKMAN COULTER (RBD45164)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 16-10-2022[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.3	5.3	10.6	10.47	0.0280	0.19	0	0.1	0.0070	-0.79
RBC x10 <sup>6</sup> /µl	1	4.31	4.29	8.6	8.75	0.0070	-0.75	0.02	0.04	0.0020	-0.54
Hb g/dl	1	12.5	12.4	24.9	25.4	0.0200	-0.75	0.1	0.1	0.0070	0.00
HCT%	1	39.4	39.3	78.7	78.2	0.1260	0.13	0.1	0.4	0.0230	-0.81
MCV-fl	1	91.8	91.3	183.1	178.7	0.2390	0.60	0.5	0.3	0.0200	0.67
MCH-Pg	1	29.1	28.9	58	58	0.0550	0.00	0.2	0.2	0.0110	0.00
MCHC-g/dl	1	31.8	31.6	63.4	64.8	0.1120	-0.42	0.2	0.3	0.0200	-0.34
Plt. x10 <sup>3</sup> /µl	1	107	106	213	214	1.05	-0.03	1	4	0.26	-0.58
Retic %	2	3.5	2.5	6	3.1	0.08	1.00	1	0.2	0.01	5.40

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=01 , Poly=35 L=12, E=02, Mono/Promono=06 , B1=03 P.M.=09, Mye=12, Meta=18, Other=
RBC Morphology	3	Poly: 42 - 56, Myelo: 13 - 23, Meta: 8 - 17, Lympho: 4 - 8, Eos: 2 - 5, Promyelo: 1-5, nRBC/Blast/Baso/Mono: 0 - 5
Diagnosis	3	Chronic Myeloid Leukemia (Chronic Phase)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 157--B	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	373	366	83.61	91.53	7.92	3.28	8.47	5.19
<b>RBC x10<sup>6</sup>/µl</b>	1	373	373	85.25	87.4	5.9	4.56	8.85	8.04
<b>Hb g/dl</b>	1	373	373	87.67	85.52	4.83	5.63	7.5	8.85
<b>HCT%</b>	1	373	365	89.59	89.59	7.12	5.21	3.29	5.2
<b>MCV-fl</b>	1	373	364	94.51	84.89	3.57	10.44	1.92	4.67
<b>MCH-Pg</b>	1	373	364	87.09	73.35	6.87	18.41	6.04	8.24
<b>MCHC-g/dl</b>	1	373	364	92.31	87.91	4.4	6.87	3.29	5.22
<b>Plt. x10<sup>3</sup>/µl</b>	1	373	364	91.21	93.13	5.22	3.3	3.57	3.57
<b>ReticCount%</b>	2	373	273	91.21	95.6	5.49	1.47	3.3	2.93
<b>PS Assessment</b>	3	373	349	Satisfactory :90.89%, Borderline Sat. :6.97%, Unsatisfactory :2.14%					

**\*Comments:**

**1). Among Lab (EQA) : PS Diagnosis wrongly reported, remaining results acceptable**

**2). Within Lab (IQA) : RETIC result is unacceptable, may be due to random/human error.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3*SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value ( $0.3*SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----