

Beckman Coulter India Pvt. Ltd. 3rd floor, B-wing, Art Guild House Phoenix Market City LBS Road, Kurla (W) Mumbai- 400 070

Ph: 022 30805000

Date: 07th September 2022

Calibration Certificate

Analyzer Model: Beckman coulter AU480

Instrument Serial no.: 2018122416

Installed at: SYMBION VIP DIAGNOSTICS

Calibrated on: 07th Sep 2022

Next calibration due: 06th Sep 2023

Photometric Check:

Requirement: All O.D. should be less than 1.7 units.

Photometric check is OK.

Incubation temperature:

Range: 37 +/- 0.5

On display the temperature is 37.0 and is stable.

We hereby confirm that Beckman Coulter AU480 has been successfully calibrated and passed All quality check and is working in good condition.

For, Beckman Coulter India Pvt. Ltd.

A.

Anil Munjani Customer Support Engineer



Beckman Coulter India Pvt. Ltd. 3rd floor, B-wing, Art Guild House Phoenix Market City LBS Road, Kurla (W) Mumbai- 400 070

Ph: 022 30805000

Calibration Certificate

Date: 07th September 2022

This is to certify that AU480 (SR. NO. 2018122416) installed at SYMBION VIP DIAGNOSTICS, Bodakdev, Ahmedabad, is calibrated as per the recommended procedures. The system parameters as mentioned below were verified and necessary calibrations done as applicable.

System temperatures

Temperature Sensor	Temperature (°C)	Operating range (°C)
Incubator Wheel	37.00	36.50 to 37.50

Power Supply Units

Power Supply	Measured (V)	Tolerances (V)
+24V	24.05	1.2V
+12V	12.01	0.4V
+5V	5.00	0.2V
-5.2V	-5.21	N.A.
-12V	-12.01	0.4V

All the parameters are within the specified Limits.

For Beckman Coulter India Pvt. Ltd.

D.

Anil Munjani Customer Support Engineer