



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 4465

Distribution No.: 158-L

Month/Year: January/2023

Instrument ID: TW-03000981

Name &amp; Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,

Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 23-02-2023[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.26	4.2	8.46	9.2	0.0280	-1.10	0.06	0.1	0.0060	-0.36
RBC x10 <sup>6</sup> /µl	1	4.83	4.8	9.63	9.42	0.0130	0.60	0.03	0.05	0.0030	-0.45
Hb g/dl	1	13.5	13.4	26.9	26.9	0.0280	0.00	0.1	0.1	0.0080	0.00
HCT%	1	44.3	43.9	88.2	85	0.2400	0.41	0.4	0.4	0.0240	0.00
MCV-fl	1	91.7	91.5	183.2	183.2	0.4090	0.00	0.2	0.2	0.0180	0.00
MCH-Pg	1	27.9	27.8	55.7	57.2	0.0640	-0.88	0.1	0.2	0.0160	-0.45
MCHC-g/dl	1	30.5	30.4	60.9	62.8	0.1560	-0.33	0.1	0.3	0.0210	-0.67
Plt. x10 <sup>3</sup> /µl	1	250	250	500	452	1.51	1.10	0	6	0.37	-1.01
Retic %	2	2.7	2.5	5.2	20.5	0.37	-1.38	0.2	0.7	0.05	-0.64

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=03 , Poly=18 L=6, E=1, Mono/Promono=5 , B1=6 P.M.=10, Mye=24, Meta=29, Other=	Poly: 28 - 51, Myelo: 12 - 22, Meta: 10- 20, Lympho: 3- 10, Eosino: 1-4, Promyelo: 2-8, nRBC/Blast/Baso/Mono: 0 - 5		
RBC Morphology	3	NORMOCYTIC NORMOCHROMIC ANEMIA	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis		
Diagnosis	3	CHRONIC MYELOID LEUKEMIA	Chronic Myeloid Leukemia (Chronic Phase)		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 158--L	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	338	337	83.09	93.47	4.15	3.56	12.76	2.97
<b>RBC x10<sup>6</sup>/µl</b>	1	338	338	88.17	88.76	6.51	4.44	5.32	6.8
<b>Hb g/dl</b>	1	338	338	86.98	84.32	5.62	6.21	7.4	9.47
<b>HCT%</b>	1	338	336	97.62	90.77	1.79	3.57	0.59	5.66
<b>MCV-fl</b>	1	338	337	99.11	85.76	0.89	3.86	0	10.38
<b>MCH-Pg</b>	1	338	337	91.69	89.91	4.45	5.34	3.86	4.75
<b>MCHC-g/dl</b>	1	338	337	98.52	88.43	0.89	5.64	0.59	5.93
<b>Plt. x10<sup>3</sup>/µl</b>	1	338	337	95.55	88.72	3.56	5.93	0.89	5.35
<b>ReticCount%</b>	2	338	217	97.7	92.17	1.84	3.23	0.46	4.60
<b>PS Assessment</b>	3	338	212	Satisfactory :93.14%, Borderline Sat. :3.43%, Unsatisfactory :3.43%					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between“0 to ±2” are texted in green colour. Z score value between“±2 to ±3” are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme’s website www.ishtmaiiseqap.com.

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----