



## Calibrator Calibration

<u>Instrument:</u> 5 Part Haematology Analyzer	<u>Model:</u> BC -6200	<u>Serial No.:</u> TW-03000981	<u>Date of Cal.</u> 12.07.2022			
<u>Mode: OV-WB</u>	<u>Calibrator Lot No. PLUS0722</u>			<u>Exp.Date:05.08.2022</u>		
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.78	4.71	14.1	92.8	235	7.94
2.	8.64	4.63	14.0	92.3	235	8.26
3.	8.97	4.64	13.9	92.0	239	8.29
4.	8.96	4.74	14.1	92.3	233	8.49
5.	8.91	4.62	13.9	92.0	234	8.33
6.	8.85	4.64	14.1	92.1	223	8.47
Mean	8.85	4.66	14.0	92.3	233	8.30
CV( %)	1.421	1.083	0.58	0.32	2.3	2.372
New Cal. Factor (%)	102.24	98.44	96.99	100.26	109.46	
Old Cal. Factors (%)	104.18	92.94	95.85	100.47	97.95	
Old TC1 (%)						102.14
New TC1 (%)						105.92

For Rapid Diagnostic Pvt. Ltd.

G. Nagaraju  
Service Engineer

N. Mahender Reddy  
Zonal Service Manager




## Calibration Certificate

<b>Customer Name</b>	<b>Telangana Diagnostic Hub Jangaon</b>
Address	OPP. Area Hospital, JANGAON-506167
Contact Person	Mr. Srinivas. P
Contact No.	9247293093
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000981
Instrument Status	Under Warranty
Calibration Done On	12.07.2022
Next Calibration Due On	11.07.2023

**For Rapid Diagnostic Pvt. Ltd.**



**G. Nagaraju**  
Service Engineer



**N. Mahender Reddy**  
Zonal Service Manager