

**STATE REFERENCE LABORATORY (SRL),SVNGMC ,YAVATMAL**  
**Timer Calibration Verification Record Sheet**

Location: ICTC Lab, SDH RAJURA

Property No: DT 103

Page No. 28

Procedure using Timer: Inhouse verification of timer using certified standard calibrated Timer

Calibration Date: 30/06/2022

Certified timer ID: SRL/MICRO/SVNGMC/37

Due Date:29/06/2023

Verification Done by: Ms Urmila R. Chambhare, Technical Officer, SRL,SVNGMC,Yavatmal


Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference
1	300	299	-1	600	599	-1	900	899	-1
2	300	300	0	600	601	1	900	900	0
3	300	300	0	600	601	1	900	899	-1
4	300	300	0	600	601	1	900	900	0
5	300	300	0	600	601	1	900	900	0
6	300	299	-1	600	600	0	900	900	0
7	300	299	-1	600	600	0	900	900	0
8	300	300	0	600	600	0	900	900	0
9	300	300	0	600	600	0	900	900	0
10	300	299	-1	600	600	0	900	899	-1
Mean	300	299.6	-0.4	600	600.3	0.3	900	901	1
CV	-	0.52	-	-	0.67	-	-	899.9	-0.1
Accuracy	-	0.172362	-	-	0.112435	-	-	0.63	-
	-	-0.133333	-	-	0.05	-	-	0.070281	-
	-	-	-	-	-	-	-	-0.011111	-

Accept/ Reject: Accepted since variation (%CV) is within 2% of all three times set

Final Interpretation: Timer is in good working condition

Operation Taken: -

Remark: -

  
 Urmila R. Chambhare  
 Technical Officer  
 SRL (Microbiology)  
 Shri V.N.G.M.C. Yavatmal