

**State Reference Laboratory
I. C. T. C.
Department of Microbiology
L. T. M. Medical College and General Hospital Mumbai-22**

Certificate of calibration

Name of ICTC: *Rawali camp Maternity Home*

Name of the Equipment: Thermometer

Date of calibration: *17/06/2022*

Serial No. / Model No.: *Mextech*

Next Calibration due on: *16/06/2023*

Identification No.: *MDACS/RCMH/ICTC/TM/01*

Acceptable range: 2 to 8 degree Celsius ($^{\circ}\text{C}$)

Point of Calibration: 2°C , 4°C , 8°C

Tolerance (If Applicable): $\pm 2^{\circ}\text{C}$

Location of Calibration: At Site

Condition of Receipt Item:

Method of calibration: In- house calibration done using Calibrated Thermometer 2021-22

Traceability Certificate no: IDEMI/CC/ECL/0556/20-21

Calibration Result:

Calibration Points ($^{\circ}\text{C}$)	UUC Reading ($^{\circ}\text{C}$)			Standard Reading ($^{\circ}\text{C}$)	Mean
	Reading 1	Reading 2	Reading 3		
2	<i>2</i>	<i>2.1</i>	<i>1.9</i>	2	2
4	<i>4</i>	<i>4</i>	<i>4</i>	4	4
8	<i>8</i>	<i>7.9</i>	<i>8.1</i>	8	8

Acceptable Variation $\pm 2^{\circ}\text{C}$ of the Mean Value

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

Rehmat
17/06/2022

Technical Officer, MDACS

intellogo


PPTCT
LG



**DO NOT STORE
FOOD AND DRINK**

MDACS/RCMH/ICTC/REF.01

Name of Equipment - PPTCT
Model No - G2A05M2110
Ser No - G2A31US04582
Location - ODD No - 205
Equipment No -
MDACS/RCMH/ICTC/REF.01



**DO NOT STORE
FOOD AND DRINK**