

**State Reference Laboratory
I. C. T. C.
Department of Microbiology
L. T. M. Medical College and General Hospital Mumbai-22**

Certificate of calibration

Name of ICTC: *Rawali camp Maternity Home*

Name of the Equipment: Thermometer

Date of calibration: *17/06/2022*

Serial No. / Model No.: *Mextech*

Next Calibration due on: *16/06/2023*

Identification No.: *MDACS/RCMH/ICTC/TM/03*

Acceptable range: 2 to 8 degree Celsius (°C)

Point of Calibration: 2°C, 4°C, 8°C

Tolerance (If Applicable): $\pm 2^\circ\text{C}$

Location of Calibration: At Site

Condition of Receipt Item:

Method of calibration: In-house calibration done using Calibrated Thermometer 2021-22

Traceability Certificate no: IDEMI/CC/ECL/0556/20-21

Calibration Result:

Calibration Points (°C)	UUC Reading (°C)			Standard Reading (°C)	Mean
	Reading 1	Reading 2	Reading 3		
2	<i>2</i>	<i>2</i>	<i>2</i>	<i>2</i>	<i>2</i>
4	<i>4.1</i>	<i>4</i>	<i>3.9</i>	<i>4</i>	<i>4</i>
8	<i>8</i>	<i>8</i>	<i>8</i>	<i>8</i>	<i>8</i>

Acceptable Variation $\pm 2^\circ\text{C}$ of the Mean Value

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

Robert
17/06/2022

Technical Officer, MDACS

59%



**DO NOT STORE
FOOD AND DRINK**

Name of Equipment - IIR
(Lentho-A Transporter)
Rental No - 3281 - 3291
Ser. No - 511021105000320
Location - OPD 710-210
Equipment No -
MDACS/RCMH/ICTC/ILR-01

MDACS/RCMH/ICTC/ILR.01

VOLTAS A TATA Product

