

**State Reference Laboratory
I. C. T. C.
Department of Microbiology
L. T. M. Medical College and General Hospital Mumbai-22**

Certificate of calibration

Name of the ICTC: *Raunali camp Maternity Home.*

Name of the Equipment: Centrifuge

Date of calibration: *17/06/2022*

Serial. No. / Model No.: *Remi/CCLC - 1091*

Next Calibration due on: *18/06/2023*

Identification No.: *MDALS/RUMH/ICTC/Cen/01*

Acceptable range: 2000 rpm to 3000 rpm

Point of Calibration: 2500 rpm & 3000 rpm

Tolerance (If Applicable): $\pm 5\%$

Location of Calibration: At Site

Condition of Receipt Item: Good

Method of calibration: In- house calibration done using Calibrated Tachometer 2021-22

Traceability Certificate no: IDEMI/CC/ECL/0568/2020-21

Calibration Result:

Calibration Points (rpm)	UUC Reading (rpm)			Standard Reading (rpm)	Mean
	Reading 1	Reading 2	Reading 3		
2500	<i>2500</i>	<i>2510</i>	<i>2490</i>	<i>2500</i>	<i>2500</i>
3000	<i>3050</i>	<i>2950</i>	<i>3000</i>	<i>3000</i>	<i>3000</i>

Acceptable Variation $\pm 5\%$ of the Mean Value

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

Ms. Anita A. Bharadwaj
17/06/2022
Technical officer, MDALS.
SRL - LTMMCL&GH, Sion
Mumbai 400 022.



Name of Equipment - Centrifuge
Model No - Remi/CCHC.
Serial No - 1091
Location - OPD No - 20.
Equipment No.
MDACS/RCMH/ICTC/CEN/01.

MDACS/RCMH/ICTC/CEN/01

REMI

RCMH/ICTC/CEN/01

