


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221201A/N-101	Page	: 1 of 1	
CALIBRATED ON.	: 01 / 12 / 2022	Ref. Std.	: ISO: 8655-6	
CALIBRATION DUE.	: 31 / 05 / 2023	C.P. No.	: ACS/CP/08	
SRF No. & DATE	: 221201A/N Dt. 01.12.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC34272200001049F Dt. 05.12.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 6, IGM Hospital Lane, (East Side of IGM Hospital), Rabindra Palli, Agartala			
ITEM CALIBRATED	: Micro Pipette			
	Make	: Thermo Scientific	Range	: (1 to 10) µl
	L. Count	:	ID No.	: RCL/HEM/PIP-01
	Location	: Haematology	S. No.	: RW19997
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. { 40% to 60% }				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023
RESULTS : Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
10		10	10.08	(-) 0.08

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.02 µl at coverage factor k=1.96.

*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

