



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221201A/N-102	Page	: 1 of 1
------------------------	-----------------	-------------	----------

CALIBRATED ON.	: 01 / 12 / 2022	Ref. Std.	: ISO: 8655-6
CALIBRATION DUE.	: 31 / 05 / 2023	C.P. No.	: ACS/CP/08

SRF No. & DATE	: 221201A/N Dt. 01.12.2022	Phy. Inspection	: O.K.
---------------------------	----------------------------	------------------------	--------

ULR No. & ISSUE DATE	: CC342722000001050F Dt. 05.12.2022	Cal. At	: Site / Lab
---------------------------------	-------------------------------------	----------------	--------------

CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. 6, IGM Hospital Lane, (East Side of IGM Hospital), Rabindra Palli, Agartala		
-----------------------	--	--	--

ITEM CALIBRATED	: Micro Pipette			
	Make	: Thermo Scientific	Range	: (20 to 200) µl
	L. Count	:	ID No.	: RCL/HEM/PIP-02
	Location	: Haematology	S. No.	: SW22995

ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C}	R.H. {40% to 60%}
-------------------------------	---------------------	-------------------

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		20	20.11	(-) 0.11
200		100	100.32	(-) 0.32
		200	200.64	(-) 0.64

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.03 µl at coverage factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....