


# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b> : 221201A/N-107	<b>Page</b> : 1 of 1			
<b>CALIBRATED ON.</b> : 01 / 12 / 2022	<b>Ref. Std.</b> : DKD R-5-7			
<b>CALIBRATION DUE.</b> : 30 / 11 / 2023	<b>C.P. No.</b> : ACS/CP/11			
<b>SRF No. &amp; DATE</b> : 221201A/N Dt. 01.12.2022	<b>Phy. Inspection</b> : O.K.			
<b>ULR No. &amp; ISSUE DATE</b> : CC342722000001055F Dt. 05.12.2022	<b>Cal. At</b> : Site / Lab			
<b>CALIBRATED FOR</b> : M/s Redcliffe Lifetech Pvt. Ltd. 6, IGM Hospital Lane, (East Side of IGM Hospital), Rabindra Palli, Agartala				
<b>ITEM CALIBRATED</b> : Digital Thermo Hygrometer				
<b>Make</b> : Digital Series	<b>Range</b> : (0 to 50)°C, (10-99)%			
<b>L. Count</b> :	<b>ID No.</b> : RCL/HEM/HGM-01			
<b>Location</b> : Haematology	<b>S. No.</b> : N.A			
<b>ENVIRONMENT CONDITION:</b> Temp. { 25°C ± 3°C } R.H. { 40% to 60% }				
<b>STANDARD INSTRUMENTS USED :</b>				
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>	<b>DUE DATE</b>
RH & Temp. Sensor with Indicator	ACS/RHTIS/01/ 20365872	DVG	CC210122000001272F	15/07/2023
<b>RESULTS</b> : Thermal Discipline				
<b>* U. C. INSTT. READS</b>	<b>STANDARD READING</b>		<b>ERROR</b>	
(°C)	15.04	15.01	0.03	
	25.04	24.98	0.06	
	40.08	39.97	0.09	
(%)	30.04	29.98	0.06	
	60.04	59.97	0.07	
	90.06	89.97	0.09	

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.8°C, 1.9% at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

