


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO. : 221201A/N-112	Page : 1 of 1			
CALIBRATED ON. : 02 / 12 / 2022	Ref. Std. : CG-11			
CALIBRATION DUE. : 01 / 12 / 2023	C.P. No. : ACS/CP/07			
SRF No. & DATE : 221201A/N Dt. 01.12.2022	Phy. Inspection : O.K.			
ULR No. & ISSUE DATE : CC342722000001060F Dt. 05.12.2022	Cal. At : Site / Lab			
CALIBRATED FOR : M/s Redcliffe Lifetech Pvt. Ltd. 6, IGM Hospital Lane, (East Side of IGM Hospital), Rabindra Palli, Agartala				
ITEM CALIBRATED : Digital Temperature Indicator				
Make : RoHS	Range : (-50 to 110) °C			
L. Count :	ID No. : RCL/BIO/TEMP-01			
Location : Clinical Biochemistry	S. No. : N.A			
ENVIRONMENT CONDITION: Temp. { 25°C ± 3°C } R.H. {40% to 60%}				
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	09/11/2023
RESULTS : ET Discipline				
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)		
2.02	2	0.02		
4.02	4	0.02		
6.04	6	0.04		
8.06	8	0.06		
(average of readings)				

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

... End of Calibration Certificate....

