



Telangana Diagnostics

Department of Health, Medical & Family Welfare
Government of Telangana



DEPARTMENT OF MICROBIOLOGY:

Respected sir/Madam

The PT report for CMC MICRO EQUAS which we have uploaded in the application is the latest report we got from them which was dispatched on 15.11. 2022. We have sent our results but not yet received the report, we have sent an email requesting the report. We will share the report as soon as we receive it. I request to consider the 15.11.2022 report as this is the latest report which we have received.



PC-1033

111th IAMM EQAS Microbiology: Bacteriology/ Serology
CMC MICRO EQAS

Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu
Email: eqas@cmcvellore.ac.in, Twitter: @microeqas, Phone: +91-416-2282588



JULY 2022

111th EQAS EVALUATION REPORT

MEMBER ID:

M 1 2 0 9

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination

Peer group (n) = 229

Marks Obtained: 10/10 (100%)

Please refer the attached evaluation format for details on the criteria for evaluation

	Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z' score	Max Marks	Your Score
SE1	CRP	Negative	Not reported	Negative	Not Applicable					2	2
SE2	CRP	Positive	48	Positive	31.0011	19.1529	6 to 192	1.6328	0.9	2	2
SE3	CRP	Positive	24	Positive	20.3592	12.0561	6 to 192	1.0525	0.3	2	2

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination

Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

	Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z' score	Max Marks	Your Score
SE1	RF	Negative	Not reported	Negative	Not Applicable					2	2
SE2	RF	Negative	Not reported	Negative	Not Applicable					2	2
SE3	RF	Positive	16	Negative	NOT EVALUATED*						

*As >30% participants scored <70% on this exercise, it has been excluded from the final evaluation.

Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 1 2 0 9


SE1	SE2	SE3	Marks obtained	
4	4	2	10	100%
4	4	2	Maximum marks = 10	



Dr. Rani Diana Sahni
Scientific Co-ordinator



Dr. John A Jude Prakash
Quality Manager



Dr. V. Balaji
PT Co-ordinator

Report Dispatch Date: 15.11.2022

***** End of Report *****