

NON-CONFORMANCE FORM
(Use one form for each parameter)

OPL/LAB/FR 06

DATE: 10/2/23

ISSUE (Description): Urea EOAS Out

INVESTIGATION: QC result, calibration investigated.

ROOT CAUSE:

CORRECTIVE ACTION: Urea Rerun with EOAS sample.
Result attached

PREVENTIVE ACTION (Where applicable):

CONTROLLED

Auth :

QA Function

REVIEWED BY: Jyoel
10/2/23

LAB DIRECTOR:

Prepared by: DR. Jyoti	Reviewed by: DR. Rector	Authorized by: Mr. Gaurav
<u>Jyoel</u>		

Result Reprint

Report Type : Patients

Order #	Sample ID	Test	Result Unit	Flag	Curve #	Result Date
1	EQAS 01	UREA	30.8 mg/dL		11995	10-Feb-2023 15:51:20

EQAS Urea = 30.96

Return value = 30.80

Difference = 0.16 (within -1.00 to +1.00) Reference
EQAS Report

NC closed on 10/2/2023

Speed
10/2/23