



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 2398

Distribution No.: 157-F

Month/Year: August/2022

Instrument ID: ERBA H-360, 3 PART (K10012116054)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 28-10-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	4.98	4.89	9.87	10.5	0.0360	-0.75	0.09	0.1	0.0070	-0.08
RBC x10 ⁶ /µl	1	4.45	4.42	8.87	8.43	0.0090	1.95	0.03	0.04	0.0030	-0.27
Hb g/dl	1	13.9	13.8	27.7	28.4	0.0300	-0.91	0.1	0.1	0.0080	0.00
HCT%	1	45.5	45	90.5	85.2	0.1910	1.07	0.5	0.4	0.0260	0.27
MCV-fl	1	102.4	101.9	204.3	202.6	0.3720	0.18	0.5	0.3	0.0240	0.54
MCH-Pg	1	31.4	31.3	62.7	67.4	0.0810	-2.44	0.1	0.3	0.0200	-0.67
MCHC-g/dl	1	30.7	30.6	61.3	66.2	0.1610	-1.16	0.1	0.3	0.0190	-0.67
Plt. x10 ³ /µl	1	149	149	298	273	1.81	0.54	0	5	0.32	-1.12
Retic %	2	3	2	5	11.4	0.22	-1.14	1	0.49	0.03	0.86

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=02 , Poly=02 L=04, E=00, Mono/Promono=00 , B1=02 P.M.=90, Mye=00, Meta=00, Other=
RBC Morphology	3	ANISOCYTOSIS++, MICROCYTIC+, HYPOCHROMIC+, NRBCS SEEN
Diagnosis	3	S/O: ACUTE PROMYELOCYTIC LEUKEMIA

Blast: 30-85, Poly: 2-8, Lympho: 4-10, Myelo: 0-7, Mono: 1-10.5, nRBC/Promyelo/Meta/Eos: 0-5

Predominantly: Normocytic/Normochromic; Moderate: Microcytosis, Hypochromia; Mild: Anisocytosis, Macrocytosis

Acute Myeloid Leukemia (AML)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 157--F	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	288	288	82.29	86.11	4.17	7.64	13.54	6.25
RBC x10⁶/µl	1	288	288	86.46	93.06	6.25	3.47	7.29	3.47
Hb g/dl	1	288	288	87.5	87.5	5.9	6.6	6.6	5.9
HCT%	1	288	287	91.99	90.59	3.83	5.92	4.18	3.49
MCV-fl	1	288	287	91.64	94.08	5.57	1.05	2.79	4.87
MCH-Pg	1	288	287	83.97	89.55	7.32	6.62	8.71	3.83
MCHC-g/dl	1	288	287	93.03	86.76	4.18	5.23	2.79	8.01
Plt. x10³/µl	1	288	287	89.2	88.85	6.27	5.57	4.53	5.58
ReticCount%	2	288	251	95.62	91.24	3.98	7.17	0.4	1.59
PS Assessment	3	288	255	Satisfactory :94.45%, Borderline Sat. :3.12%, Unsatisfactory :2.43%					

***Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Precision acceptable.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----