



Add : Flat No.302, Third Floor, Krishna Pride Apartment,
Sadguru Nagar, Pathardi Gaon, Nashik-422 010.
Mob.: +91 9028646172, +91 9028777244 (J) 0253 4034044
E-mail : rktechnologies99@gmail.com
Website : www.rktechcalibration.com

Calibration of Electro technical,
Thermal, Pressure, Mechanical,
Force & Hardness parameters.

Validation of All types of
Temperature & Humidity parameters.

CALIBRATION CERTIFICATE

Calibration Item	Haematology Analyser	Certificate No	RK/21-22/12-01
------------------	----------------------	----------------	----------------

Date of Receipt	Date of Calibration	Next Recommended Due Date	Certificate Issue Date	Page No
17 December 2022	17 December 2022	16 December 2023	19 December 2022	01 of 02

I. Customer Name & Address	M/S. DLS Diagnostic Laboratory Services Bajaj Nagar, More Chowk Waluj Dist : Aurangabad
----------------------------	--

II. Description of Item Under Calibration :			
Instrument ID No	DLS/CE-01		
Make /Model	Erba	Location	LAB

III Environment Condition :			
Temperature	20 to 35 °C	Procedure of Calibration	RK /MECH/ 09
Relative Humidity	45 % to 75 % RH	Reference Standard Used	ASTM D-2240-05
Location Of Calibration	At Site	Conditioned of Receipt Item	OK

IV. Detail of Reference Standard used for calibration (Traceable To National / International Standard)					
Instrument Name	ID No	Traceability (Cert No)	Date of Calibration	Valid upto	Traceability
Power Analyzer	RK-STD-50	NEC/22-23/AA/006-1	11 April 2022	10 April 2023	NABL, CC-2248

V: Calibration Result :

VISUAL TEST

Sr No	Parameter	Result
1	Power Cord, Cables Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B, BF, CF)	B
4	Equipment Class (I,II,III)	I

B-Body Type, BF-Body Float Type, CF- Cardiac Float Type
I- Properly Earthed, II-Double Insulation, III-With Extra Low Safety Voltage

FUNCTIONAL TEST

Sr No	Parameter	Observed	Tolerance	Result
1	Voltage between Live and Neutral (Vln)	234.85	(210-240 V)	Pass
2	Voltage between Live and Earth (Vle)	237.95	(210-240 V)	Pass
3	Voltage between Neutral and Earth (Vne)	3.4	(0-5 V)	Pass
4	Enclosure Current (IL)	0.79	(0-10µA)	Pass
5	Leakage Current (IL)	119.40	(0-500µA)	Pass
6	Patient Leakage Current (PL)	0.76	(BF-0 -100µA) (CF-0-100µA)	Pass





RK Technologies

Calibration and Validation Services

NABL ACCREDITED LABORATORY
AS PER STANDARD 17025 : 2017

Add : Flat No.302, Third Floor, Krishna Pride Apartment,
Sadguru Nagar, Pathardi Gaon, Nashik-422 010.

Mob.: +91 9028646172, +91 9028777244 ☎ 0253 4034044

E-mail : rktechnologies99@gmail.com

Website : www.rktechcalibration.com


Calibration of Electro technical,
Thermal, Pressure, Mechanical,
Force & Hardness parameters.


Validation of All types of
Temperature & Humidity parameters.


Date of Receipt	Date of Calibration	Next Recommended Due Date	Certificate Issue Date	Page No
17 December 2022	17 December 2022	16 December 2023	19 December 2022	02 of 02

VI : Note:

- 1) UUC stands for Unit Under Calibration.
- 2) Next calibration date (1 Year) mentioned in the certificate is given as per customer request
- 3) This certificate refers only to the particular item submitted for calibration
- 4) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "RK Technologies, Nashik".
- 5) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.


Calibrated By
Mr. Yogesh Berad
Calibration Engineer


Review & Approved By
Mr. Rahul Kasture
Technical Manager



***** End of Report *****