

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Erythrocyte (RBC) Count	Mindray BC-5150	1.15	mill/cu.mm	-2.14
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

BHUMEE PATHOLOGY LAB PRIVATE LIMITED (595)

Outlier And Analyte Summary Report
Result Details For Cycle No 220412 and Sample No 03


Report Date : 25/01/2023

Instrument : Mindray BC-5150

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
! Erythrocyte (RBC) Count	mill/cu.mm	1.15	--	-2.14	-5.88
@ Total Leucocytes (WBC) count	/c.mm	1420.00	--	-0.14	-1.33
@ Haemoglobin	g/dL	3.10	--	-0.04	-5.66
@ MCV (Mean Corpuscular Volume)	fL	105.90	--	-0.54	3.19
@ MCH (Mean Corpuscular Hb)	pg	26.50	--	1.00	-0.15
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	25.10	--	1.58	-1.02
@ Platelet Count	1000/ μ L	41.00	--	0.82	0.34
@ Hematocrit	%	12.20	--	-0.83	-8.59
@ RDW CV%	%	25.00	--	1.68	13.93

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

End of Report