



We Put You First,
& Keep You Ahead !!

OM SALES CORPORATION

Certificate of Calibration

Customer Name : Bhume Pathology Laboratory Pvt.Ltd., Ahmednagar

Model : Automated Heamatology Analyzer BC -5150

Serial No. : SR - 79002353

Calibration Done Date : 05.11.2022

Next Calibration Due Date On or Before : 04.11.2023

Lab In-Charge :

This is to certify that the above mentioned product has been verified of calibration for CBC 5 parameter (**WBC, RBC, HGB, HCT, and PLT**) according to the standard Procedure provided by Mindray Corporation, China.

The reference instrument used for value assignment are managed by the traceability Mindray in Mindray Corporation and these are traceable to the international standards, Such as **ICSH**

Calibration at site performed by
Engineer Name : Yash Ghanekar
Designation : Service Engineer

Om Sales Corporation
Location : Pune.

