

ADACS (RCMH) TM / 01  
HSETIN  
INSTRUMENTS  
DT-1045  
Extra Loud Sound



Electronic Count Down/Up Timer



RESET

State Reference Laboratory  
I. C. T. C.  
Department of Microbiology  
L. T. M. Medical College and General Hospital Mumbai-22

**Certificate of calibration**

Name of ICTC: Rawali camp Maternity Home

Name of the Equipment: Timer (Digital) Date of calibration: 17/06/2022

Serial No. / Model No.: ERMA-DT 1045 Next Calibration due on: 16/06/2023

Identification No.: MDACS/RCMH/ICTC/DT/01

Acceptable range: 0 - 60 Mins

Point of Calibration: 5 mins, 10 mins, 20 mins

Tolerance (If Applicable):  $\pm 2\%$

Location of Calibration: At Site

Condition of Receipt Item:

Method of calibration: In- house calibration done using Calibrated Digital stop watch 2021-22

Traceability Certificate no: IDEMI/CC/ECL/0562/20-21

Calibration Result:

Calibration Points (Minutes)	UUC Reading (Minutes)			Standard Reading (Minutes)	Mean
	Reading 1	Reading 2	Reading 3		
5	5	5.1	4.9	5	5
10	10.1	10.0	9.9	10	10
20	20	20.1	19.9	20	20

Acceptable Variation  $\pm 2\%$  of the Mean Value

UUC: Unit Under Calibration

Remarks: Instrument has passed / ~~failed~~ the Acceptance Criteria

Checked & Approved by:

*Reent*  
Ms. Anita A. Bharadwaj  
Technical officer, MDACS  
SEL-LTMMELH, Son  
Mumbai 400022.