

## Calibration Report

Sample ID 1  
Lot number CX473

Name ABX MINOCAL

Exp. date 09/05/2022  
Modified on

Coefficient	WBC	RBC	HGB	HCT	PLT	MPV
New	1.080	0.890	0.941	1.059	1.005	1.020
Current	1.109	0.851	0.970	1.078	0.866	1.000
Target	9.00	4.51	12.8	37.4	260	10.3
Mean	9.24	4.31	13.2	36.4	224	10.1
CV(%)	1.66	1.26	0.23	1.25	1.90	1.72

Number of calibration run selected for coefficient calculation (minimum 5) 8/10

Sel.	Run Time	WBC (10 <sup>3</sup> /μL)	RBC (10 <sup>6</sup> /μL)	HGB (g/dL)	HCT (%)	PLT (10 <sup>3</sup> /μL)	MPV (μm <sup>3</sup> )
	08/29/2022 02:04:36 PM	9.33 h	4.34 l	13.4 h	36.8	245 l	10.6
✓	08/29/2022 02:05:57 PM	9.16	4.23 l	13.2 h	35.6 l	223 l	10.5
✓	08/29/2022 02:07:22 PM	9.00	4.32 l	13.2 h	36.7	224 l	9.9
✓	08/29/2022 02:08:59 PM	9.12	4.27 l	13.2 h	36.2 l	230 l	10.0
	08/29/2022 02:10:35 PM	9.33 h	4.28 l	13.2 h	36.2 l	230 l	10.5
✓	08/29/2022 02:12:13 PM	9.39 h	4.35 l	13.2 h	36.8	219 l	10.1
✓	08/29/2022 02:13:53 PM	9.16	4.25 l	13.2 h	36.1 l	222 l	10.2
✓	08/29/2022 02:15:31 PM	9.30 h	4.38 l	13.3 h	37.0	229 l	10.0
✓	08/29/2022 02:17:12 PM	9.45 h	4.32 l	13.2 h	36.3 l	219 l	10.2
✓	08/29/2022 02:18:54 PM	9.34 h	4.35 l	13.2 h	36.8	227 l	10.2