

BHUMEE PATHOLOGY LAB PRIVATE LIMITED (595)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 220412 and Sample No 01

Report Date : 25/01/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

BHUMEE PATHOLOGY LAB PRIVATE LIMITED (595)

Outlier And Analyte Summary Report


Result Details For Cycle No 220412 and Sample No 01

Report Date : 25/01/2023

Instrument : Biorad D-10

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ HbA1c	%	6.30	6.47	-0.86	-0.34

Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.	Total Parameters	1
	Not Evaluated Parameters	0
	Evaluated Parameters	1
	Outlier Parameters (X)	0
	EQAS Score HbA1c	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A,
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

End of Report